

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JPCL01

Lab Code: CHEM

SAS No.: Q1487 SDG NO.: Q1487

Lab File ID: BF141834.D

DFTPP Injection Date: 03/05/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35.5
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	47.3
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	13.5
442	Greater than 50% of mass 198	84.3
443	15.0 - 24.0% of mass 442	16.7 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141835.D	03/05/2025	10:09
DN-B-42MS	Q1487-01MS	BF141844.D	03/05/2025	15:21
DN-B-42MSD	Q1487-01MSD	BF141845.D	03/05/2025	15:51
DN-B-42	Q1487-01	BF141846.D	03/05/2025	16:21