

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2025

Project # N/A

Chemtech Project # Q1488

Test Name: TCLP BNA

A. Number of Samples and Date of Receipt:

12 Solid samples were received on 03/04/2025.

3 Water samples were received on 03/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The samples were analyzed on instrument BNA_G using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GG. The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for ENV-101-SB01 [2,4,6-Tribromophenol - 124%], ENV-101-SB01MS [2,4,6-Tribromophenol - 122%], ENV-101-SB01MSD [2,4,6-Tribromophenol - 116%], ENV-101-SB02 [2,4,6-Tribromophenol - 113%], ENV-102-SB01 [2,4,6-Tribromophenol - 119%], ENV-102-SB02 [2,4,6-Tribromophenol - 116%], ENV-104-SB02 [2,4 and 6-Tribromophenol - 113%]. These surrogate did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF022725.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The % RSD is greater than 20% in the Initial Calibration (8270-BG030525.M) for 2,4-Dinitrotoluen this compounds is passing on Linear Regression.

The Continuous Calibration File ID BF141882.D met the requirements except for 2,4-Dinitrotoluene, 2,4,6-Tribromophenol and Nitrobenzene-d5 .The associate samples have no positive hit for these compounds therefore no corrective action was required.

Pentachlorophenol is failing high and associate sample having hit but below CRQL therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____