

### Laboratory Control Sample Summary

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q1488

**Project:** Amtrak Sawtooth Bridges 2025

**Run No.:** LB134896

| Analyte             | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | LB134896BS |            |        |                 |            |                 |                     |               |
| Hexavalent Chromium | mg/L       | 0.5        | 0.50   |                 | 101        | 1               | 90-111              | 03/04/2025    |

### Laboratory Control Sample Summary

**Client:** Portal Partners Tri-Venture

**SDG No.:** Q1488

**Project:** Amtrak Sawtooth Bridges 2025

**Run No.:** LB134940

| Analyte             | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|---------------------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID           | PB167023BS |            |        |                 |            |                 |                     |               |
| Hexavalent Chromium | mg/Kg      | 20         | 20.1   |                 | 101        | 1               | 84-110              | 03/07/2025    |