

**Metals**  
- 4 -  
**INTERFERENCE CHECK SAMPLE**

|                                            |                          |
|--------------------------------------------|--------------------------|
| <b>Client:</b> Portal Partners Tri-Venture | <b>SDG No.:</b> Q1488    |
| <b>Contract:</b> PORT06                    | <b>Lab Code:</b> CHEM    |
| <b>Case No.:</b> Q1488                     | <b>SAS No.:</b> Q1488    |
| <b>ICS Source:</b> EPA                     | <b>Instrument ID:</b> P4 |

| Sample ID | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSA01    | Arsenic  | 6.46           |                    |               | -20                    | 20                      | 03/11/2025       | 11:06            | LB134987      |
|           | Barium   | 5.32           | 6.0                | 89            | -94                    | 106                     | 03/11/2025       | 11:06            | LB134987      |
|           | Cadmium  | -2.31          | 1.0                | 231           | -5                     | 7                       | 03/11/2025       | 11:06            | LB134987      |
|           | Chromium | 58.2           | 52.0               | 112           | 42                     | 62                      | 03/11/2025       | 11:06            | LB134987      |
|           | Lead     | 3.81           |                    |               | -12                    | 12                      | 03/11/2025       | 11:06            | LB134987      |
|           | Selenium | -16.3          |                    |               | -20                    | 20                      | 03/11/2025       | 11:06            | LB134987      |
|           | Silver   | -0.94          |                    |               | -10                    | 10                      | 03/11/2025       | 11:06            | LB134987      |
| ICSAB01   | Arsenic  | 114            | 104                | 110           | 88.4                   | 120                     | 03/11/2025       | 11:11            | LB134987      |
|           | Barium   | 487            | 537                | 91            | 437                    | 637                     | 03/11/2025       | 11:11            | LB134987      |
|           | Cadmium  | 1020           | 972                | 105           | 826                    | 1120                    | 03/11/2025       | 11:11            | LB134987      |
|           | Chromium | 567            | 542                | 105           | 460                    | 624                     | 03/11/2025       | 11:11            | LB134987      |
|           | Lead     | 55.7           | 49.0               | 114           | 37                     | 61                      | 03/11/2025       | 11:11            | LB134987      |
|           | Selenium | 32.4           | 46.0               | 70            | 26                     | 66                      | 03/11/2025       | 11:11            | LB134987      |
|           | Silver   | 220            | 201                | 110           | 170                    | 232                     | 03/11/2025       | 11:11            | LB134987      |
| ICSA01    | Arsenic  | 5.78           |                    |               | -20                    | 20                      | 03/12/2025       | 11:09            | LB135011      |
|           | Barium   | 1.43           | 6.0                | 24            | -94                    | 106                     | 03/12/2025       | 11:09            | LB135011      |
|           | Cadmium  | -3.50          | 1.0                | 350           | -5                     | 7                       | 03/12/2025       | 11:09            | LB135011      |
|           | Chromium | 60.0           | 52.0               | 115           | 42                     | 62                      | 03/12/2025       | 11:09            | LB135011      |
|           | Lead     | 5.99           |                    |               | -12                    | 12                      | 03/12/2025       | 11:09            | LB135011      |
|           | Selenium | -15.8          |                    |               | -20                    | 20                      | 03/12/2025       | 11:09            | LB135011      |
|           | Silver   | 0.57           |                    |               | -10                    | 10                      | 03/12/2025       | 11:09            | LB135011      |
| ICSAB01   | Arsenic  | 116            | 104                | 112           | 88.4                   | 120                     | 03/12/2025       | 11:14            | LB135011      |
|           | Barium   | 454            | 537                | 84            | 437                    | 637                     | 03/12/2025       | 11:14            | LB135011      |
|           | Cadmium  | 1000           | 972                | 103           | 826                    | 1120                    | 03/12/2025       | 11:14            | LB135011      |
|           | Chromium | 595            | 542                | 110           | 460                    | 624                     | 03/12/2025       | 11:14            | LB135011      |
|           | Lead     | 52.3           | 49.0               | 107           | 37                     | 61                      | 03/12/2025       | 11:14            | LB135011      |
|           | Selenium | 34.2           | 46.0               | 74            | 26                     | 66                      | 03/12/2025       | 11:14            | LB135011      |
|           | Silver   | 226            | 201                | 112           | 170                    | 232                     | 03/12/2025       | 11:14            | LB135011      |
| ICSA01    | Arsenic  | 4.32           |                    |               | -20                    | 20                      | 03/13/2025       | 12:25            | LB135035      |
|           | Barium   | 4.72           | 6.0                | 79            | -94                    | 106                     | 03/13/2025       | 12:25            | LB135035      |
|           | Cadmium  | -3.11          | 1.0                | 311           | -5                     | 7                       | 03/13/2025       | 12:25            | LB135035      |
|           | Chromium | 60.0           | 52.0               | 115           | 42                     | 62                      | 03/13/2025       | 12:25            | LB135035      |
|           | Lead     | 4.72           |                    |               | -12                    | 12                      | 03/13/2025       | 12:25            | LB135035      |
|           | Selenium | -15.2          |                    |               | -20                    | 20                      | 03/13/2025       | 12:25            | LB135035      |
|           | Silver   | -0.88          |                    |               | -10                    | 10                      | 03/13/2025       | 12:25            | LB135035      |
| ICSAB01   | Arsenic  | 113            | 104                | 109           | 88.4                   | 120                     | 03/13/2025       | 12:30            | LB135035      |
|           | Barium   | 485            | 537                | 90            | 437                    | 637                     | 03/13/2025       | 12:30            | LB135035      |
|           | Cadmium  | 1000           | 972                | 103           | 826                    | 1120                    | 03/13/2025       | 12:30            | LB135035      |
|           | Chromium | 564            | 542                | 104           | 460                    | 624                     | 03/13/2025       | 12:30            | LB135035      |
|           | Lead     | 54.2           | 49.0               | 111           | 37                     | 61                      | 03/13/2025       | 12:30            | LB135035      |
|           | Selenium | 43.0           | 46.0               | 94            | 26                     | 66                      | 03/13/2025       | 12:30            | LB135035      |

**Metals**

- 4 -

**INTERFERENCE CHECK SAMPLE**

**Client:** Portal Partners Tri-Venture **SDG No.:** Q1488  
**Contract:** PORT06 **Lab Code:** CHEM **Case No.:** Q1488 **SAS No.:** Q1488  
**ICS Source:** EPA **Instrument ID:** P4

| Sample ID | Analyte | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSAB01   | Silver  | 218            | 201                | 108           | 170                    | 232                     | 03/13/2025       | 12:30            | LB135035      |