

Continuing Calibration Report for SequenceID : FC030525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068368.D

Aliphatic EPH	40236785.000	360.000	3.102	21.556	111768.847	122175.127	8.518
Aliphatic C9-C12	7078387.000	60.000	3.102	6.366	117973.117	141035.502	16.352
Aliphatic C16-C21	6684413.000	60.000	9.750	13.103	111406.883	131630.895	15.364
Aliphatic C21-C28	9380589.000	80.000	13.104	16.755	117257.363	122691.322	4.429
Aliphatic C28-C40	12521412.000	120.000	16.756	21.556	104345.100	102846.462	-1.457
Aliphatic C12-C16	4571984.000	40.000	6.367	9.749	114299.600	136654.515	16.359