

Continuing Calibration Report for SequenceID : FC030525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068376.D

Aliphatic EPH	44765920.000	360.000	3.102	21.556	124349.778	122175.127	-1.780
Aliphatic C9-C12	8592043.000	60.000	3.102	6.366	143200.717	141035.502	-1.535
Aliphatic C16-C21	8029066.000	60.000	9.750	13.103	133817.767	131630.895	-1.661
Aliphatic C21-C28	10022524.000	80.000	13.104	16.755	125281.550	122691.322	-2.111
Aliphatic C28-C40	12601397.000	120.000	16.756	21.556	105011.642	102846.462	-2.105
Aliphatic C12-C16	5520890.000	40.000	6.367	9.749	138022.250	136654.515	-1.001