

Continuing Calibration Report for SequenceID : FD030525AR

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FD049158.D

|                  |              |         |        |        |            |            |        |
|------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aromatic C10-C12 | 6556391.000  | 40.000  | 4.075  | 5.793  | 163909.775 | 159020.606 | -3.075 |
| Aromatic C12-C16 | 10059952.000 | 60.000  | 5.794  | 8.395  | 167665.867 | 160935.866 | -4.182 |
| Aromatic C16-C21 | 12177228.000 | 80.000  | 8.396  | 12.660 | 152215.350 | 153359.309 | 0.746  |
| Aromatic C21-C36 | 25536741.000 | 180.000 | 12.661 | 18.072 | 141870.783 | 143114.753 | 0.869  |
| Aromatic EPH     | 54330312.000 | 360.000 | 4.075  | 18.072 | 150917.533 | 150128.823 | -0.525 |