

Continuing Calibration Report for SequenceID : FD030525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049166.D

Aromatic C10-C12	6523459.000	40.000	4.075	5.793	163086.475	159020.606	-2.557
Aromatic C12-C16	10110953.000	60.000	5.794	8.395	168515.883	160935.866	-4.710
Aromatic C16-C21	12497804.000	80.000	8.396	12.660	156222.550	153359.309	-1.867
Aromatic C21-C36	25202288.000	180.000	12.661	18.072	140012.711	143114.753	2.168
Aromatic EPH	54334504.000	360.000	4.075	18.072	150929.178	150128.823	-0.533