

Continuing Calibration Report for SequenceID : FE031025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052707.D

Aliphatic C9-C12	9658611.000	60.000	3.151	6.796	160976.850	155948.436	-3.224
Aliphatic C12-C16	6261558.000	40.000	6.797	10.250	156538.950	152098.489	-2.919
Aliphatic C16-C21	8993508.000	60.000	10.251	13.630	149891.800	146548.085	-2.282
Aliphatic C21-C28	11838283.000	80.000	13.631	17.305	147978.538	142451.851	-3.880
Aliphatic C28-C40	16154870.000	120.000	17.306	22.234	134623.917	127589.108	-5.514
Aliphatic EPH	52906830.000	360.000	3.151	22.234	146963.417	141501.588	-3.860