

Continuing Calibration Report for SequenceID : FF031025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015628.D

Aromatic C10-C12	5303513.000	40.000	4.505	6.377	132587.825	128361.211	-3.293
Aromatic C12-C16	8424119.000	60.000	6.378	9.069	140401.983	136800.309	-2.633
Aromatic C16-C21	10513836.000	80.000	9.070	13.390	131422.950	133038.999	1.215
Aromatic C21-C36	21153413.000	180.000	13.391	18.845	117518.961	130055.838	9.640
Aromatic EPH	45394881.000	360.000	4.505	18.845	126096.892	131654.549	4.221