

Continuing Calibration Report for SequenceID : FF031025AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015645.D

Aromatic C10-C12	5366584.000	40.000	4.505	6.377	134164.600	128361.211	-4.521
Aromatic C12-C16	8486524.000	60.000	6.378	9.069	141442.067	136800.309	-3.393
Aromatic C16-C21	10783294.000	80.000	9.070	13.390	134791.175	133038.999	-1.317
Aromatic C21-C36	22388226.000	180.000	13.391	18.845	124379.033	130055.838	4.365
Aromatic EPH	47024628.000	360.000	4.505	18.845	130623.967	131654.549	0.783