

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1488 **SAS No :** Q1488 **SDG No:** Q1488
Sample No : PB166976BS **Datafile:** FC068370.D
Client ID : PB166976BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	152	51		(40-140)
Aliphatic C12-C16	200	152	76		(40-140)
Aliphatic C16-C21	300	262	87		(40-140)
Aliphatic C21-C28	400	362	90		(40-140)
Aliphatic C28-C40	600	596	99		(40-140)