

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1488 **SAS No :** Q1488 **SDG No:** Q1488
Sample No : PB166976BSD **Datafile:** FC068371.D
Client ID : PB166976BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	152	51		0.0657 (40-140)	50
Aliphatic C12-C16	200	152	76		0 (40-140)	50
Aliphatic C16-C21	300	263	88		0.41 (40-140)	50
Aliphatic C21-C28	400	362	90		0 (40-140)	50
Aliphatic C28-C40	600	597	100		0.2 (40-140)	50