

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1488 **SAS No :** Q1488 **SDG No:** Q1488
Sample No : PB166976BS **Datafile:** FD049160.D
Client ID : PB166976BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	149	74		(40-140)
Aromatic C12-C16	300	274	91		(40-140)
Aromatic C16-C21	500	525	105		(40-140)
Aromatic C21-C36	800	822	102		(40-140)