

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1488 **SAS No :** Q1488 **SDG No:** Q1488
Sample No : PB166976BSD **Datafile:** FD049161.D
Client ID : PB166976BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	200	151	76		1.3 (40-140)	50
Aromatic C12-C16	300	280	93		2.2 (40-140)	50
Aromatic C16-C21	500	534	106		1.7 (40-140)	50
Aromatic C21-C36	800	835	104		2.3 (40-140)	50