

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109424.D	20 Feb 2025 16:09		YP/AJ	Ok
2	I.BLK	I.BLK	PO109425.D	20 Feb 2025 16:28		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO109426.D	20 Feb 2025 16:46		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO109427.D	20 Feb 2025 17:04		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO109428.D	20 Feb 2025 17:23		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO109429.D	20 Feb 2025 17:41		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO109430.D	20 Feb 2025 17:59		YP/AJ	Ok
8	AR1221ICC500	AR1221ICC500	PO109431.D	20 Feb 2025 18:18		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO109432.D	20 Feb 2025 18:36		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO109433.D	20 Feb 2025 18:55		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO109434.D	20 Feb 2025 19:13		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO109435.D	20 Feb 2025 19:31		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO109436.D	20 Feb 2025 19:50		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO109437.D	20 Feb 2025 20:08		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO109438.D	20 Feb 2025 20:26		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO109439.D	20 Feb 2025 20:45		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO109440.D	20 Feb 2025 21:03		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO109441.D	20 Feb 2025 21:21		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO022025

Review By	yogesh	Review On	2/21/2025 8:09:53 AM
Supervise By	Ankita	Supervise On	2/21/2025 9:30:37 AM
SubDirectory	PO022025	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO109442.D	20 Feb 2025 21:40		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO109443.D	20 Feb 2025 21:58		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO109444.D	20 Feb 2025 22:17		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO109445.D	20 Feb 2025 22:35		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO109446.D	20 Feb 2025 22:53		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO109447.D	20 Feb 2025 23:12		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO109448.D	20 Feb 2025 23:30		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO109449.D	20 Feb 2025 23:48		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO109450.D	21 Feb 2025 00:07		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO109451.D	21 Feb 2025 00:25		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO109452.D	21 Feb 2025 00:43		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO109453.D	21 Feb 2025 01:02		YP/AJ	Ok
31	PO022025ICV500	ICVPO022025	PO109454.D	21 Feb 2025 01:20		YP/AJ	Ok
32	AR1242ICV500	ICVPO022025AR1242	PO109455.D	21 Feb 2025 01:38		YP/AJ	Ok
33	AR1248ICV500	ICVPO022025AR1248	PO109456.D	21 Feb 2025 01:57		YP/AJ	Ok
34	AR1254ICV500	ICVPO022025AR1254	PO109457.D	21 Feb 2025 02:15		YP/AJ	Ok
35	AR1268ICV500	ICVPO022025AR1268	PO109458.D	21 Feb 2025 02:34		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO109647.D	06 Mar 2025 09:28		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO109648.D	06 Mar 2025 10:05		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO109649.D	06 Mar 2025 10:24		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO109650.D	06 Mar 2025 10:42		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO109651.D	06 Mar 2025 11:00		YP/AJ	Ok
6	I.BLK	I.BLK	PO109652.D	06 Mar 2025 11:20		YP/AJ	Ok
7	PB167003BL	PB167003BL	PO109653.D	06 Mar 2025 12:29		YP/AJ	Ok
8	PB167003BS	PB167003BS	PO109654.D	06 Mar 2025 12:47		YP/AJ	Ok
9	Q1498-01	4A-4B-4C-1928 BLD-W	PO109655.D	06 Mar 2025 13:06		YP/AJ	Ok,M
10	Q1498-02	5A-5B-5C-1985 BLDG	PO109656.D	06 Mar 2025 13:24		YP/AJ	Ok,M
11	Q1498-03	6A-6B-6C-1985	PO109657.D	06 Mar 2025 13:43		YP/AJ	Ok
12	Q1498-04	7A-7B-7C-ROOF-5	PO109658.D	06 Mar 2025 14:01		YP/AJ	Ok,M
13	Q1498-05	8A-8B-8C-ROOF-5	PO109659.D	06 Mar 2025 14:20		YP/AJ	Ok
14	Q1498-06	9A-9B-9C-ROOF-6	PO109660.D	06 Mar 2025 14:37		YP/AJ	Ok
15	Q1498-07	10A-10B-10C-ROOF-6	PO109661.D	06 Mar 2025 14:55		YP/AJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PO109662.D	06 Mar 2025 16:11		YP/AJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PO109663.D	06 Mar 2025 16:47		YP/AJ	Ok
18	AR1248CCC500	AR1248CCC500	PO109664.D	06 Mar 2025 17:06		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO109665.D	06 Mar 2025 17:24		YP/AJ	Ok
20	I.BLK	I.BLK	PO109666.D	06 Mar 2025 18:01		YP/AJ	Ok
21	PB167009BL	PB167009BL	PO109667.D	06 Mar 2025 18:19		YP/AJ	Ok
22	PB167009BS	PB167009BS	PO109668.D	06 Mar 2025 18:38		YP/AJ	Ok
23	PB167009BSD	PB167009BSD	PO109669.D	06 Mar 2025 18:56		YP/AJ	Ok
24	Q1488-13	ENV-102-GW01	PO109670.D	06 Mar 2025 19:15		YP/AJ	Ok
25	Q1488-14	ENV-104-GW01	PO109671.D	06 Mar 2025 19:33		YP/AJ	Ok
26	Q1494-01	PURGE-WATER	PO109672.D	06 Mar 2025 19:51		YP/AJ	Ok
27	Q1437-07	K084-13B	PO109673.D	06 Mar 2025 20:10	AR1254 + AR1268 Hit, Need 5x Dilution	YP/AJ	Dilution
28	Q1437-10	K084-14B	PO109674.D	06 Mar 2025 20:28	AR1254 + AR1268 Hit	YP/AJ	Ok,M
29	Q1438-03	K084-15C	PO109675.D	06 Mar 2025 20:47	AR1254 + AR1268 Hit, Need 20x Dilution	YP/AJ	Dilution
30	Q1438-05	K084-16C	PO109676.D	06 Mar 2025 21:04	AR1254 + AR1268 Hit, Need 10x Dilution	YP/AJ	Dilution
31	AR1660CCC500	AR1660CCC500	PO109677.D	06 Mar 2025 22:19		YP/AJ	Ok,M
32	AR1242CCC500	AR1242CCC500	PO109678.D	06 Mar 2025 22:56		YP/AJ	Ok
33	AR1248CCC500	AR1248CCC500	PO109679.D	06 Mar 2025 23:14		YP/AJ	Ok
34	AR1254CCC500	AR1254CCC500	PO109680.D	06 Mar 2025 23:33		YP/AJ	Ok
35	I.BLK	I.BLK	PO109681.D	07 Mar 2025 00:09		YP/AJ	Ok
36	PB167022BL	PB167022BL	PO109682.D	07 Mar 2025 00:28		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO030625

Review By	yogesh	Review On	3/6/2025 10:53:50 AM
Supervise By	mohammad	Supervise On	3/8/2025 4:31:13 AM
SubDirectory	PO030625	HP Acquire Method	HP Processing Method PO022025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167022BS	PB167022BS	PO109683.D	07 Mar 2025 00:46		YP/AJ	Ok
38	Q1433-03	K084-1C	PO109684.D	07 Mar 2025 01:05	AR1254 + AR1268 Hit	YP/AJ	Ok,M
39	Q1433-06	K084-2C	PO109685.D	07 Mar 2025 01:23	AR1254 + AR1268 Hit, Need 2x Dilution	YP/AJ	Dilution
40	Q1433-09	K084-3C	PO109686.D	07 Mar 2025 01:41	AR1254 + AR1268 Hit, Need 4x dilution	YP/AJ	Dilution
41	Q1434-02	K084-4C	PO109687.D	07 Mar 2025 02:00	AR1254 + AR1268 Hit, Need 5x dilution	YP/AJ	Dilution
42	Q1434-02MS	K084-4CMS	PO109688.D	07 Mar 2025 02:18	AR1660 Recovery Fai	YP/AJ	Ok,M
43	Q1434-02MSD	K084-4CMSD	PO109689.D	07 Mar 2025 02:37		YP/AJ	Ok,M
44	Q1434-05	K084-5C	PO109690.D	07 Mar 2025 02:55	AR1254 + AR1268 Hit, Need 5x dilution	YP/AJ	Dilution
45	Q1434-07	K084-6C	PO109691.D	07 Mar 2025 03:13	AR1254 + AR1268 Hit, need 4x dilution, DCB High in both column	YP/AJ	Dilution
46	AR1660CCC500	AR1660CCC500	PO109692.D	07 Mar 2025 04:29		YP/AJ	Ok,M
47	AR1242CCC500	AR1242CCC500	PO109693.D	07 Mar 2025 05:05		YP/AJ	Ok
48	AR1248CCC500	AR1248CCC500	PO109694.D	07 Mar 2025 05:24		YP/AJ	Ok
49	AR1254CCC500	AR1254CCC500	PO109695.D	07 Mar 2025 05:42		YP/AJ	Ok,M
50	I.BLK	I.BLK	PO109696.D	07 Mar 2025 06:19		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069994.D	24 Feb 2025 14:27		YPIAJ	Ok
2	I.BLK	I.BLK	PP069995.D	24 Feb 2025 14:43		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP069996.D	24 Feb 2025 14:59		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP069997.D	24 Feb 2025 15:15		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP069998.D	24 Feb 2025 15:32		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP069999.D	24 Feb 2025 15:48		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP070000.D	24 Feb 2025 16:04		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070001.D	24 Feb 2025 16:20		YPIAJ	Ok,M
9	AR1232ICC500	AR1232ICC500	PP070002.D	24 Feb 2025 16:37		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP070003.D	24 Feb 2025 16:53		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PP070004.D	24 Feb 2025 17:09		YPIAJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP070005.D	24 Feb 2025 17:25		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PP070006.D	24 Feb 2025 17:42		YPIAJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP070007.D	24 Feb 2025 17:58		YPIAJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP070008.D	24 Feb 2025 18:14		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PP070009.D	24 Feb 2025 18:30		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PP070010.D	24 Feb 2025 18:46		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PP070011.D	24 Feb 2025 19:03		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP022425

Review By	yogesh	Review On	2/25/2025 9:00:36 AM
Supervise By	Ankita	Supervise On	2/25/2025 9:48:29 AM
SubDirectory	PP022425	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP070012.D	24 Feb 2025 19:19		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP070013.D	24 Feb 2025 19:35		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP070014.D	24 Feb 2025 19:51		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP070015.D	24 Feb 2025 20:08		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP070016.D	24 Feb 2025 20:24		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP070017.D	24 Feb 2025 20:40		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP070018.D	24 Feb 2025 20:56		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP070019.D	24 Feb 2025 21:12		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP070020.D	24 Feb 2025 21:29		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP070021.D	24 Feb 2025 21:45		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP070022.D	24 Feb 2025 22:01		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP070023.D	24 Feb 2025 22:17		YPIAJ	Ok,M
31	PP022425ICV500	ICVPP022425	PP070024.D	24 Feb 2025 22:34		YPIAJ	Ok
32	AR1242ICV500	ICVPP022425AR1242	PP070025.D	24 Feb 2025 22:50		YPIAJ	Ok
33	AR1248ICV500	ICVPP022425AR1248	PP070026.D	24 Feb 2025 23:06		YPIAJ	Ok
34	AR1254ICV500	ICVPP022425AR1254	PP070027.D	24 Feb 2025 23:22		YPIAJ	Ok
35	AR1268ICV500	ICVPP022425AR1268	PP070028.D	24 Feb 2025 23:38		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By	mohammad	Supervise On	3/7/2025 1:28:32 AM
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070250.D	05 Mar 2025 08:31		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070251.D	05 Mar 2025 08:47		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070252.D	05 Mar 2025 09:04		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070253.D	05 Mar 2025 09:20		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070254.D	05 Mar 2025 09:36		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP070255.D	05 Mar 2025 09:52		YPIAJ	Ok
7	Q1476-01	TRE-25-0014	PP070256.D	05 Mar 2025 10:09	AR1242 4 peak report, DCB LOW both column	YPIAJ	ReRun
8	Q1476-01RE	TRE-25-0014RE	PP070257.D	05 Mar 2025 10:25	AR1242 4 peak report, DCB LOW both column	YPIAJ	Confirms
9	Q1473-02RE	BUR-1293RE	PP070258.D	05 Mar 2025 11:00	DCB not detected	YPIAJ	Confirms
10	AR1660CCC500	AR1660CCC500	PP070259.D	05 Mar 2025 11:49		YPIAJ	Ok
11	AR1242CCC500	AR1242CCC500	PP070260.D	05 Mar 2025 12:05		YPIAJ	Ok,M
12	AR1248CCC500	AR1248CCC500	PP070261.D	05 Mar 2025 12:21		YPIAJ	Ok,M
13	AR1254CCC500	AR1254CCC500	PP070262.D	05 Mar 2025 12:38		YPIAJ	Ok,M
14	I.BLK	I.BLK	PP070263.D	05 Mar 2025 12:54		YPIAJ	Ok
15	PB166985BL	PB166985BL	PP070264.D	05 Mar 2025 14:58		YPIAJ	Ok
16	PB166985BS	PB166985BS	PP070265.D	05 Mar 2025 15:14		YPIAJ	Ok
17	Q1481-01	WC-K1311	PP070266.D	05 Mar 2025 15:31		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By	mohammad	Supervise On	3/7/2025 1:28:32 AM
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1482-01	OR-03-030425	PP070267.D	05 Mar 2025 15:47		YPIAJ	Ok,M
19	Q1484-01	TR-OTR-COMP-01	PP070268.D	05 Mar 2025 16:03		YPIAJ	Ok
20	Q1485-01	DN-B-41	PP070269.D	05 Mar 2025 16:19		YPIAJ	Ok
21	Q1486-01	DN-B-40	PP070270.D	05 Mar 2025 16:36		YPIAJ	Ok
22	Q1487-01	DN-B-42	PP070271.D	05 Mar 2025 16:52		YPIAJ	Ok
23	Q1488-01	ENV-101-SB01	PP070272.D	05 Mar 2025 17:08		YPIAJ	Ok
24	Q1488-03	ENV-101-SB02	PP070273.D	05 Mar 2025 17:25		YPIAJ	Ok
25	AR1660CCC500	AR1660CCC500	PP070274.D	05 Mar 2025 18:13		YPIAJ	Ok
26	AR1242CCC500	AR1242CCC500	PP070275.D	05 Mar 2025 18:30		YPIAJ	Ok,M
27	AR1248CCC500	AR1248CCC500	PP070276.D	05 Mar 2025 18:46		YPIAJ	Ok,M
28	AR1254CCC500	AR1254CCC500	PP070277.D	05 Mar 2025 19:02		YPIAJ	Ok,M
29	I.BLK	I.BLK	PP070278.D	05 Mar 2025 19:19		YPIAJ	Ok
30	Q1488-03MS	ENV-101-SB02MS	PP070279.D	05 Mar 2025 19:35		YPIAJ	Ok
31	Q1488-03MSD	ENV-101-SB02MSD	PP070280.D	05 Mar 2025 19:51		YPIAJ	Ok,M
32	Q1488-05	ENV-102-SB01	PP070281.D	05 Mar 2025 20:08		YPIAJ	Ok
33	Q1488-07	ENV-102-SB02	PP070282.D	05 Mar 2025 20:24		YPIAJ	Ok
34	Q1488-09	ENV-104-SB01	PP070283.D	05 Mar 2025 20:40		YPIAJ	Ok
35	Q1488-11	ENV-104-SB02	PP070284.D	05 Mar 2025 20:56		YPIAJ	Ok
36	Q1489-01	TP-1	PP070285.D	05 Mar 2025 21:13		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP030525

Review By	yogesh	Review On	3/5/2025 11:09:53 AM
Supervise By	mohammad	Supervise On	3/7/2025 1:28:32 AM
SubDirectory	PP030525	HP Acquire Method	HP Processing Method PP022425
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q1489-05	TP-2	PP070286.D	05 Mar 2025 21:29		YPIAJ	Ok
38	Q1489-09	TP-3	PP070287.D	05 Mar 2025 21:45		YPIAJ	Ok
39	Q1489-13	TP-4	PP070288.D	05 Mar 2025 22:02		YPIAJ	Ok
40	AR1660CCC500	AR1660CCC500	PP070289.D	05 Mar 2025 22:50		YPIAJ	Ok
41	AR1242CCC500	AR1242CCC500	PP070290.D	05 Mar 2025 23:07		YPIAJ	Ok,M
42	AR1248CCC500	AR1248CCC500	PP070291.D	05 Mar 2025 23:23		YPIAJ	Ok,M
43	AR1254CCC500	AR1254CCC500	PP070292.D	05 Mar 2025 23:39		YPIAJ	Ok,M
44	I.BLK	I.BLK	PP070293.D	05 Mar 2025 23:56		YPIAJ	Ok

M : Manual Integration