

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

Instrument ID: ECD_S Calibration Date(s): 02/21/2025 02/21/2025
Calibration Times: 19:56 21:32

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029234.D</u>	CF 500 =	<u>PS029235.D</u>		
CF 750 =		<u>PS029236.D</u>	CF 1000 =	<u>PS029237.D</u>	CF 1500 =	<u>PS029238.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	28468600000	25326000000	23950000000	23359600000	21830500000	24586900000	10
2,4-D	5578960000	4794680000	4576300000	4498750000	4303460000	4750430000	10
2,4-DCAA	4834500000	4175340000	3953240000	3904380000	3735430000	4120580000	10

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

Instrument ID: ECD_S

Calibration Date(s): 02/21/2025 02/21/2025

Calibration Times: 19:56 21:32

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029234.D</u>	CF 500 =	<u>PS029235.D</u>		
CF 750 =		<u>PS029236.D</u>	CF 1000 =	<u>PS029237.D</u>	CF 1500 =	<u>PS029238.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	9297750000	9072730000	8996050000	9011120000	8805850000	9036700000	2
2,4-D	1363000000	1257020000	1239960000	1254990000	1234500000	1269900000	4
2,4-DCAA	1014780000	934728000	917942000	923871000	910342000	940333000	5