

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031225\
 Data File : PS029417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 18:15
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/13/2025
 Supervised By : Ankita Jodhani 03/13/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 02:32:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022125.M
 Quant Title : 8080.M
 QLast Update : Sat Feb 22 01:02:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.154	7.619	3285.5E6	704.9E6	797.345m	749.627
Target Compounds							
1) T	Dalapon	2.599	2.642	3776.3E6	1403.3E6	684.081	662.829
2) T	3,5-DICHL...	6.338	6.594	4347.4E6	997.0E6	743.002	704.024
3) T	4-Nitroph...	6.955	7.151	1605.7E6	540.8E6	686.163	678.061
5) T	DICAMBA	7.337	7.811	13324.0E6	3807.5E6	767.246	738.890m
6) T	MCP	7.517	7.918	988.4E6	164.6E6	80.192m	70.388
7) T	MCPA	7.665	8.157	1288.1E6	227.8E6	76.013	71.423
8) T	DICHLORPROP	8.037	8.521	3422.1E6	927.0E6	744.506	724.927
9) T	2,4-D	8.267	8.846	3493.7E6	914.3E6	735.449	719.998
10) T	Pentachlo...	8.558	9.362	45135.2E6	17687.1E6	773.443	768.582
11) T	2,4,5-TP ...	9.131	9.739	18906.4E6	6879.9E6	768.962	761.325
12) T	2,4,5-T	9.423	10.154	19473.9E6	6105.7E6	764.020	743.665
13) T	2,4-DB	9.992	10.716	3006.5E6	587.9E6	720.044	767.918m
14) T	DINOSEB	11.184	11.090	12348.9E6	4301.8E6	742.531m	709.687m
15) T	Picloram	11.001	12.164	21035.8E6	7885.2E6	688.717	681.971
16) T	DCPA	11.482	12.124	23198.8E6	9171.5E6	773.204	774.846

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS031225\
Data File : PS029417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 18:15
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/13/2025
Supervised By : Ankita Jodhani 03/13/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 02:32:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS022125.M
Quant Title : 8080.M
QLast Update : Sat Feb 22 01:02:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

