

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

Continuing Calib Date: 03/13/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 20:04 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.05	9.06	8.96	9.16	0.01
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.91	4.92	4.82	5.02	0.01
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

Continuing Calib Date: 03/13/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 20:04 Initial Calibration Time(s): 10:35 11:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.90	7.91	7.81	8.01	0.01
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
gamma-BHC (Lindane)	3.60	3.60	3.50	3.70	0.00
Heptachlor	3.94	3.94	3.84	4.04	0.00
Heptachlor epoxide	4.72	4.73	4.63	4.83	0.01
Endrin	5.63	5.63	5.53	5.73	0.00
Methoxychlor	6.60	6.61	6.51	6.71	0.01