

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094653.D Time Analyzed: 19:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.050	8.956	9.156	57.460	50.000	14.9
Endrin	6.570	6.474	6.674	46.520	50.000	-7.0
gamma-BHC (Lindane)	4.325	4.227	4.427	51.810	50.000	3.6
Heptachlor	4.912	4.815	5.015	48.640	50.000	-2.7
Heptachlor epoxide	5.681	5.583	5.783	50.640	50.000	1.3
Methoxychlor	7.497	7.400	7.600	50.870	50.000	1.7
Tetrachloro-m-xylene	3.537	3.438	3.638	52.660	50.000	5.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/12/2025

Lab Sample No.: PSTDCCC050 Data File : PL094653.D Time Analyzed: 19:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.903	7.807	8.007	56.290	50.000	12.6
Endrin	5.631	5.534	5.734	51.880	50.000	3.8
gamma-BHC (Lindane)	3.602	3.504	3.704	55.610	50.000	11.2
Heptachlor	3.940	3.842	4.042	52.170	50.000	4.3
Heptachlor epoxide	4.721	4.625	4.825	54.580	50.000	9.2
Methoxychlor	6.604	6.507	6.707	50.690	50.000	1.4
Tetrachloro-m-xylene	2.771	2.672	2.872	55.740	50.000	11.5