

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/13/2025

Lab Sample No.: PSTDCCC050 Data File : PL094695.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.048	8.956	9.156	53.870	50.000	7.7
Endrin	6.568	6.474	6.674	46.050	50.000	-7.9
gamma-BHC (Lindane)	4.324	4.227	4.427	47.020	50.000	-6.0
Heptachlor	4.911	4.815	5.015	45.500	50.000	-9.0
Heptachlor epoxide	5.679	5.583	5.783	45.670	50.000	-8.7
Methoxychlor	7.496	7.400	7.600	54.200	50.000	8.4
Tetrachloro-m-xylene	3.535	3.438	3.638	46.690	50.000	-6.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL05 Date Analyzed: 03/13/2025

Lab Sample No.: PSTDCCC050 Data File : PL094695.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.902	7.807	8.007	51.700	50.000	3.4
Endrin	5.629	5.534	5.734	48.760	50.000	-2.5
gamma-BHC (Lindane)	3.601	3.504	3.704	49.290	50.000	-1.4
Heptachlor	3.939	3.842	4.042	48.060	50.000	-3.9
Heptachlor epoxide	4.720	4.625	4.825	49.010	50.000	-2.0
Methoxychlor	6.602	6.507	6.707	52.890	50.000	5.8
Tetrachloro-m-xylene	2.770	2.672	2.872	49.280	50.000	-1.4