

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: **PORT06**

Lab Code: **CHEM** Case No.: **Q1488** SAS No.: **Q1488** SDG NO.: **Q1488**

GC Column: **ZB-MR1** ID: **0.32** (mm) Initi. Calib. Date(s): **03/11/2025** **03/11/2025**

Client Sample No. (PEM): **PEM - PL094621.D** Date Analyzed: **03/12/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **08:51**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.054	8.950	9.150	22.150	20.000	10.8
Tetrachloro-m-xylene	3.539	3.490	3.590	20.850	20.000	4.3
alpha-BHC	3.994	3.940	4.040	10.750	10.000	7.5
beta-BHC	4.526	4.480	4.580	10.570	10.000	5.7
gamma-BHC (Lindane)	4.327	4.280	4.380	10.700	10.000	7.0
Endrin	6.573	6.500	6.640	41.560	50.000	-16.9
4,4'-DDT	7.025	6.950	7.100	87.210	100.000	-12.8
Methoxychlor	7.500	7.430	7.570	217.020	250.000	-13.2

GC Column: **ZB-MR2** ID: **0.32** (mm) Initi. Calib. Date(s): **03/11/2025** **03/11/2025**

Client Sample No. (PEM): **PEM - PL094621.D** Date Analyzed: **03/12/2025**

Lab Sample No.(PEM): **PEM** Time Analyzed: **08:51**

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.906	7.810	8.010	20.130	20.000	0.7
Tetrachloro-m-xylene	2.772	2.720	2.820	20.240	20.000	1.2
alpha-BHC	3.274	3.220	3.320	9.640	10.000	-3.6
beta-BHC	3.904	3.850	3.950	10.800	10.000	8.0
gamma-BHC (Lindane)	3.604	3.550	3.650	9.650	10.000	-3.5
Endrin	5.634	5.560	5.700	44.940	50.000	-10.1
4,4'-DDT	6.032	5.960	6.100	96.750	100.000	-3.3
Methoxychlor	6.606	6.540	6.680	223.560	250.000	-10.6