

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
 Data File : PL094640.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2025 15:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 05:34:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.771	138.9E6	184.6E6	49.079	51.707
28)	SA Decachlor...	9.053	7.904	108.0E6	214.7E6	51.252	53.143
Target Compounds							
2)	A alpha-BHC	3.993	3.273	203.1E6	283.8E6	48.920	52.639
3)	MA gamma-BHC...	4.326	3.603	194.2E6	270.0E6	48.680	52.543
4)	MA Heptachlor	4.913	3.940	184.7E6	268.4E6	47.572	50.946
5)	MB Aldrin	5.255	4.219	179.0E6	255.1E6	48.473	52.307
6)	B beta-BHC	4.524	3.903	87177853	116.0E6	47.245	52.236
7)	B delta-BHC	4.771	4.131	186.5E6	265.4E6	47.901	53.068
8)	B Heptachlo...	5.682	4.722	162.5E6	237.5E6	48.578	51.877
9)	A Endosulfan I	6.067	5.091	148.6E6	210.1E6	48.385	47.873
10)	B gamma-Chl...	5.938	4.972	163.6E6	251.4E6	48.556	52.064
11)	B alpha-Chl...	6.016	5.035	159.8E6	246.7E6	48.485	51.696
12)	B 4,4'-DDE	6.191	5.224	148.0E6	259.0E6	50.298	55.725
13)	MA Dieldrin	6.342	5.356	154.7E6	268.2E6	48.387	55.276
14)	MA Endrin	6.572	5.631	135.1E6	220.4E6	48.722	50.501
15)	B Endosulfa...	6.792	5.926	130.8E6	223.3E6	48.189	51.588
16)	A 4,4'-DDD	6.708	5.779	112.6E6	198.0E6	52.004	55.062
17)	MA 4,4'-DDT	7.021	6.029	113.4E6	201.5E6	47.667	49.984
18)	B Endrin al...	6.922	6.106	101.2E6	167.7E6	47.920	49.843
19)	B Endosulfa...	7.157	6.328	118.8E6	211.7E6	48.867	51.965
20)	A Methoxychlor	7.499	6.604	59449549	106.6E6	49.662	50.261
21)	B Endrin ke...	7.642	6.833	131.8E6	254.2E6	49.875	53.268
22)	Mirex	8.114	7.013	101.3E6	191.1E6	49.010	50.359
24)	Chlordane-2	5.255f	4.358	179.0E6	1261236	1182.870	7.142 #
25)	Chlordane-3	5.938	4.972	163.6E6	251.4E6	327.321	472.530 #
26)	Chlordane-4	6.016	5.035	159.8E6	246.7E6	271.345	473.976 #
27)	Chlordane-5	0.000	5.926	0	223.3E6	N.D.	1163.946 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031225\
Data File : PL094640.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2025 15:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 05:34:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: -0.001 min
Response: 138927334
Conc: 49.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: -0.001 min
Response: 184555158
Conc: 51.71 ng/ml




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: -0.001 min
Response: 203132221
Conc: 48.92 ng/ml



#2 alpha-BHC


R.T.: 3.273 min
Delta R.T.: -0.002 min
Response: 283796994
Conc: 52.64 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.326 min
Delta R.T.: -0.001 min
Response: 194245543
Conc: 48.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.603 min
Delta R.T.: -0.002 min
Response: 270042292
Conc: 52.54 ng/ml




#4 Heptachlor

R.T.: 4.913 min
Delta R.T.: -0.001 min
Response: 184655395
Conc: 47.57 ng/ml



#4 Heptachlor


R.T.: 3.940 min
Delta R.T.: -0.002 min
Response: 268415894
Conc: 50.95 ng/ml



#5 Aldrin


R.T.: 5.255 min
Delta R.T.: -0.001 min
Response: 178974429
Conc: 48.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.219 min
Delta R.T.: -0.003 min
Response: 255064741
Conc: 52.31 ng/ml




#6 beta-BHC

R.T.: 4.524 min
Delta R.T.: -0.001 min
Response: 87177853
Conc: 47.25 ng/ml



#6 beta-BHC


R.T.: 3.903 min
Delta R.T.: -0.002 min
Response: 116031264
Conc: 52.24 ng/ml



#7 delta-BHC


R.T.: 4.771 min
Delta R.T.: -0.002 min
Response: 186541022
Conc: 47.90 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.131 min
Delta R.T.: -0.002 min
Response: 265446317
Conc: 53.07 ng/ml




#8 Heptachlor epoxide

R.T.: 5.682 min
Delta R.T.: -0.001 min
Response: 162503681
Conc: 48.58 ng/ml



#8 Heptachlor epoxide


R.T.: 4.722 min
Delta R.T.: -0.003 min
Response: 237521802
Conc: 51.88 ng/ml



#9 Endosulfan I


R.T.: 6.067 min
Delta R.T.: -0.002 min
Response: 148552584
Conc: 48.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.091 min
Delta R.T.: -0.003 min
Response: 210102820
Conc: 47.87 ng/ml




#10 gamma-Chlordane

R.T.: 5.938 min
Delta R.T.: -0.002 min
Response: 163601780
Conc: 48.56 ng/ml



#10 gamma-Chlordane


R.T.: 4.972 min
Delta R.T.: -0.002 min
Response: 251384851
Conc: 52.06 ng/ml



#11 alpha-Chlordane


R.T.: 6.016 min
Delta R.T.: -0.002 min
Response: 159847503
Conc: 48.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.035 min
Delta R.T.: -0.002 min
Response: 246733167
Conc: 51.70 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: -0.002 min
Response: 147980482
Conc: 50.30 ng/ml



#12 4,4'-DDE


R.T.: 5.224 min
Delta R.T.: -0.002 min
Response: 259037303
Conc: 55.72 ng/ml



#13 Dieldrin


R.T.: 6.342 min
Delta R.T.: -0.002 min
Response: 154746823
Conc: 48.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.356 min
Delta R.T.: -0.003 min
Response: 268187323
Conc: 55.28 ng/ml




#14 Endrin

R.T.: 6.572 min
Delta R.T.: -0.002 min
Response: 135060193
Conc: 48.72 ng/ml



#14 Endrin


R.T.: 5.631 min
Delta R.T.: -0.003 min
Response: 220368318
Conc: 50.50 ng/ml



#15 Endosulfan II


R.T.: 6.792 min
Delta R.T.: -0.002 min
Response: 130821538
Conc: 48.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.926 min
Delta R.T.: -0.003 min
Response: 223286820
Conc: 51.59 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: -0.002 min
Response: 112644733
Conc: 52.00 ng/ml



#16 4,4'-DDD


R.T.: 5.779 min
Delta R.T.: -0.003 min
Response: 197997598
Conc: 55.06 ng/ml



#17 4,4'-DDT


R.T.: 7.021 min
Delta R.T.: -0.002 min
Response: 113378079
Conc: 47.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.029 min
Delta R.T.: -0.002 min
Response: 201541637
Conc: 49.98 ng/ml




#18 Endrin aldehyde

R.T.: 6.922 min
Delta R.T.: -0.002 min
Response: 101159534
Conc: 47.92 ng/ml



#18 Endrin aldehyde


R.T.: 6.106 min
Delta R.T.: -0.003 min
Response: 167740594
Conc: 49.84 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.157 min
Delta R.T.: -0.002 min
Response: 118845797
Conc: 48.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.328 min
Delta R.T.: -0.002 min
Response: 211677614
Conc: 51.97 ng/ml




#20 Methoxychlor

R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 59449549
Conc: 49.66 ng/ml



#20 Methoxychlor


R.T.: 6.604 min
Delta R.T.: -0.003 min
Response: 106607770
Conc: 50.26 ng/ml



#21 Endrin ketone


R.T.: 7.642 min
Delta R.T.: -0.002 min
Response: 131832179
Conc: 49.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.833 min
Delta R.T.: -0.002 min
Response: 254227874
Conc: 53.27 ng/ml




#22 Mirex

R.T.: 8.114 min
Delta R.T.: -0.002 min
Response: 101267858
Conc: 49.01 ng/ml



#22 Mirex


R.T.: 7.013 min
Delta R.T.: -0.003 min
Response: 191129576
Conc: 50.36 ng/ml



#24 Chlordane-2


R.T.: 5.255 min
Delta R.T.: 0.025 min
Response: 178974429
Conc: 1182.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 4.358 min
Delta R.T.: 0.014 min
Response: 1261236
Conc: 7.14 ng/ml




#25 Chlordane-3

R.T.: 5.938 min
Delta R.T.: -0.002 min
Response: 163601780
Conc: 327.32 ng/ml



#25 Chlordane-3


R.T.: 4.972 min
Delta R.T.: -0.003 min
Response: 251384851
Conc: 472.53 ng/ml



#26 Chlordane-4


R.T.: 6.016 min
Delta R.T.: -0.005 min
Response: 159847503
Conc: 271.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.035 min
Delta R.T.: -0.002 min
Response: 246733167
Conc: 473.98 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.871 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.926 min
Delta R.T.: -0.007 min
Response: 223286820
Conc: 1163.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: -0.003 min
Response: 108010444
Conc: 51.25 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.904 min
Delta R.T.: -0.003 min
Response: 214665498
Conc: 53.14 ng/ml