

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG No.: Q1488

Instrument ID: MSVOA_X Calibration Date/Time: 03/06/2025 09:56

Lab File ID: VX045155.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.764	0.725		-5.11	20
1,1-Dichloroethene	0.614	0.600		-2.28	20
2-Butanone	0.555	0.564		1.62	20
Carbon Tetrachloride	0.465	0.458		-1.5	20
Chloroform	1.239	1.209		-2.42	20
Benzene	1.448	1.464		1.11	20
1,2-Dichloroethane	0.521	0.534		2.49	20
Trichloroethene	0.340	0.341		0.29	20
Tetrachloroethene	0.320	0.322		0.63	20
Chlorobenzene	1.058	1.078	0.3	1.89	20
1,2-Dichloroethane-d4	0.795	0.783		-1.51	20
Dibromofluoromethane	0.334	0.341		2.1	20
Toluene-d8	1.212	1.244		2.64	20
4-Bromofluorobenzene	0.402	0.419		4.23	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.