

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045155.D
 Acq On : 06 Mar 2025 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 07 00:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.629	0.631	-0.3	97	0.00
3 P	Chloromethane	0.770	0.723	6.1	93	0.00
4 C	Vinyl Chloride	0.764	0.725	5.1#	92	0.00
5 T	Bromomethane	0.300	0.280	6.7	92	0.00
6 T	Chloroethane	0.352	0.298	15.3	77	0.00
7 T	Trichlorofluoromethane	1.011	0.973	3.8	89	0.00
8 T	Diethyl Ether	0.396	0.363	8.3	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.616	-6.0	100	0.00
10 T	Methyl Iodide	0.726	0.750	-3.3	94	0.00
11 T	Tert butyl alcohol	0.151	0.118	21.9	85	0.02
12 CM	1,1-Dichloroethene	0.614	0.600	2.3#	93	0.00
13 T	Acrolein	0.174	0.149	14.4	83	0.00
14 T	Allyl chloride	1.094	1.096	-0.2	95	0.00
15 T	Acrylonitrile	0.404	0.398	1.5	94	0.00
16 T	Acetone	0.369	0.345	6.5	95	0.00
17 T	Carbon Disulfide	1.636	1.462	10.6	85	0.00
18 T	Methyl Acetate	0.888	0.957	-7.8	106	0.00
19 T	Methyl tert-butyl Ether	2.061	2.015	2.2	93	0.00
20 T	Methylene Chloride	0.731	0.704	3.7	97	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.610	-0.5	93	0.00
22 T	Diisopropyl ether	2.224	2.259	-1.6	95	0.00
23 T	Vinyl Acetate	1.856	1.909	-2.9	93	0.00
24 P	1,1-Dichloroethane	1.247	1.215	2.6	94	0.00
25 T	2-Butanone	0.555	0.564	-1.6	94	0.00
26 T	2,2-Dichloropropane	0.585	0.797	-36.2#	129	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.750	-0.1	95	0.00
28 T	Bromochloromethane	0.594	0.551	7.2	88	0.00
29 T	Tetrahydrofuran	0.373	0.365	2.1	94	0.00
30 C	Chloroform	1.239	1.209	2.4#	94	0.00
31 T	Cyclohexane	1.082	1.086	-0.4	95	0.00
32 T	1,1,1-Trichloroethane	1.000	0.978	2.2	92	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.783	1.5	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.334	0.341	-2.1	98	0.00
36 T	1,1-Dichloropropene	0.459	0.458	0.2	92	0.00
37 T	Ethyl Acetate	0.591	0.593	-0.3	92	0.00
38 T	Carbon Tetrachloride	0.465	0.458	1.5	92	0.00
39 T	Methylcyclohexane	0.549	0.598	-8.9	96	0.00
40 TM	Benzene	1.448	1.464	-1.1	92	0.00
41 T	Methacrylonitrile	0.319	0.332	-4.1	92	0.00
42 TM	1,2-Dichloroethane	0.521	0.534	-2.5	96	0.00
43 T	Isopropyl Acetate	0.874	0.903	-3.3	92	0.00
44 TM	Trichloroethene	0.340	0.341	-0.3	95	0.00
45 C	1,2-Dichloropropane	0.372	0.377	-1.3#	96	0.00
46 T	Dibromomethane	0.263	0.272	-3.4	95	0.00
47 T	Bromodichloromethane	0.516	0.523	-1.4	94	0.00
48 T	Methyl methacrylate	0.438	0.456	-4.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045155.D
 Acq On : 06 Mar 2025 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 07 00:36:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	93	0.00
50 S	Toluene-d8	1.212	1.244	-2.6	97	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.616	-4.9	95	0.00
52 CM	Toluene	0.849	0.893	-5.2#	94	0.00
53 T	t-1,3-Dichloropropene	0.431	0.467	-8.4	94	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.544	-8.2	96	0.00
55 T	1,1,2-Trichloroethane	0.350	0.355	-1.4	94	0.00
56 T	Ethyl methacrylate	0.532	0.557	-4.7	92	0.00
57 T	1,3-Dichloropropane	0.605	0.612	-1.2	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.284	-9.7	94	0.00
59 T	2-Hexanone	0.425	0.451	-6.1	95	0.00
60 T	Dibromochloromethane	0.366	0.378	-3.3	93	0.00
61 T	1,2-Dibromoethane	0.349	0.353	-1.1	94	0.00
62 S	4-Bromofluorobenzene	0.402	0.419	-4.2	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.320	0.322	-0.6	94	0.00
65 PM	Chlorobenzene	1.058	1.078	-1.9	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.353	-0.3	94	0.00
67 C	Ethyl Benzene	1.843	1.931	-4.8#	94	0.00
68 T	m/p-Xylenes	0.675	0.704	-4.3	92	0.00
69 T	o-Xylene	0.681	0.694	-1.9	93	0.00
70 T	Styrene	1.101	1.163	-5.6	93	0.00
71 P	Bromoform	0.266	0.266	0.0	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.903	3.979	-1.9	94	0.00
74 T	N-amyl acetate	1.831	1.898	-3.7	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.414	1.3	97	0.00
76 T	1,2,3-Trichloropropane	1.164	1.142	1.9	97	0.00
77 T	Bromobenzene	0.921	0.925	-0.4	96	0.00
78 T	n-propylbenzene	4.409	4.654	-5.6	94	0.00
79 T	2-Chlorotoluene	2.808	2.793	0.5	95	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.242	-2.7	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.352	-3.5	101	0.00
82 T	4-Chlorotoluene	3.060	3.111	-1.7	94	0.00
83 T	tert-Butylbenzene	3.242	3.270	-0.9	94	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.275	-3.1	94	0.00
85 T	sec-Butylbenzene	3.884	4.093	-5.4	95	0.00
86 T	p-Isopropyltoluene	3.139	3.371	-7.4	96	0.00
87 T	1,3-Dichlorobenzene	1.643	1.638	0.3	96	0.00
88 T	1,4-Dichlorobenzene	1.662	1.644	1.1	95	0.00
89 T	n-Butylbenzene	2.703	3.021	-11.8	98	0.00
90 T	Hexachloroethane	0.568	0.565	0.5	91	0.00
91 T	1,2-Dichlorobenzene	1.648	1.687	-2.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.286	-2.5	96	0.00
93 T	1,2,4-Trichlorobenzene	0.938	1.010	-7.7	98	0.00
94 T	Hexachlorobutadiene	0.385	0.403	-4.7	100	0.00
95 T	Naphthalene	3.605	3.838	-6.5	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045155.D
Acq On : 06 Mar 2025 09:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 07 00:36:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.054	-6.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6