

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Mahesh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133187,VP133189
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199
CCC	VP133188,VP133190
Internal Standard/PEM	
ICV/I.BLK	VP133200
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045067.D	28 Feb 2025 01:03		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045068.D	28 Feb 2025 01:27		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045069.D	28 Feb 2025 02:13		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045070.D	28 Feb 2025 02:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045071.D	28 Feb 2025 03:00		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045072.D	28 Feb 2025 03:23		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045073.D	28 Feb 2025 03:47		JC/MD	Ok,M
8	IBLK	IBLK	VX045074.D	28 Feb 2025 04:10		JC/MD	Ok
9	VSTDICV050	ICVVX022825	VX045075.D	28 Feb 2025 04:33		JC/MD	Ok,M
10	BFB	BFB	VX045076.D	28 Feb 2025 10:03		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045077.D	28 Feb 2025 10:32	V12668	JC/MD	Ok,M
12	VX0228MBL01	VX0228MBL01	VX045078.D	28 Feb 2025 11:00		JC/MD	Ok
13	VX0228WBL01	VX0228WBL01	VX045079.D	28 Feb 2025 11:23		JC/MD	Ok
14	VX0228WBS01	VX0228WBS01	VX045080.D	28 Feb 2025 11:46		JC/MD	Ok,M
15	VX0228WBSD01	VX0228WBSD01	VX045081.D	28 Feb 2025 12:13		JC/MD	Ok,M
16	Q1401-03	BP-VPB-192-GW-840-8	VX045082.D	28 Feb 2025 12:37	vial B pH<2	JC/MD	Ok
17	Q1401-06	BP-VPB-192-GW-900-9	VX045083.D	28 Feb 2025 13:00	vial B pH<2	JC/MD	Ok
18	Q1423-01	BP-VPB-192-TB-20250	VX045084.D	28 Feb 2025 13:23	vial B pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Maresh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133187,VP133189		
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199		
CCC	VP133188,VP133190		
Internal Standard/PEM			
ICV/I.BLK	VP133200		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1423-03	BP-VPB-192-EB-20250	VX045085.D	28 Feb 2025 13:47	vial B pH<2 EB	JC/MD	Ok
20	Q1403-01	Storage-Blank-SOIL-RE	VX045086.D	28 Feb 2025 14:10	vial B pH<2	JC/MD	Ok
21	Q1403-02	Storage-Blank-WATER-	VX045087.D	28 Feb 2025 14:33	vial B pH<2	JC/MD	Ok
22	Q1435-01	286107	VX045088.D	28 Feb 2025 14:57	vial A pH<2	JC/MD	Ok
23	Q1403-04	Storage-Blank-SAMPLE	VX045089.D	28 Feb 2025 15:20	vial B pH<2	JC/MD	Ok
24	Q1462-02	FB	VX045090.D	28 Feb 2025 15:44	vial A pH<2 FB	JC/MD	Ok
25	Q1469-01	Storage-Blank-SOIL-RE	VX045091.D	28 Feb 2025 16:07	vial A pH<2	JC/MD	Ok
26	Q1469-02	Storage-Blank-WATER-	VX045092.D	28 Feb 2025 16:31	vial A pH<2	JC/MD	Ok
27	Q1469-03	Storage-Blank-WATER-	VX045093.D	28 Feb 2025 16:54	vial A pH<2	JC/MD	Ok
28	Q1469-04	Storage-Blank-SAMPLE	VX045094.D	28 Feb 2025 17:17	vial A pH<2	JC/MD	Ok
29	Q1403-03	Storage-Blank-WATER-	VX045095.D	28 Feb 2025 17:41	vial B pH<2	JC/MD	Ok
30	Q1462-01	MW2	VX045096.D	28 Feb 2025 18:04	vial A pH<2	JC/MD	Ok,M
31	IBLK	IBLK	VX045097.D	28 Feb 2025 18:27		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050EC	VX045098.D	28 Feb 2025 18:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133217
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045126.D	05 Mar 2025 08:40		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045127.D	05 Mar 2025 09:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0305MBL01	VX0305MBL01	VX045128.D	05 Mar 2025 09:39		JC/MD	Ok
4	VX0305WBL01	VX0305WBL01	VX045129.D	05 Mar 2025 10:03		JC/MD	Ok
5	VX0305WBS01	VX0305WBS01	VX045130.D	05 Mar 2025 10:26		JC/MD	Ok,M
6	VX0305MBS01	VX0305MBS01	VX045131.D	05 Mar 2025 10:53		JC/MD	Ok,M
7	Q1480-17	TP-3	VX045132.D	05 Mar 2025 11:16		JC/MD	Ok
8	Q1448-04ME	P3ME	VX045133.D	05 Mar 2025 11:40		JC/MD	Ok
9	IBLK	IBLK	VX045134.D	05 Mar 2025 12:03		JC/MD	Ok
10	VX0305WBSD01	VX0305WBSD01	VX045135.D	05 Mar 2025 12:26		JC/MD	Ok,M
11	Q1478-04DL	IDW-AQ-DRUM-616-02	VX045136.D	05 Mar 2025 12:50	vial B pH<2	JC/MD	Ok
12	Q1478-06DL	IDW-AQ-DRUM-614-02	VX045137.D	05 Mar 2025 13:13	vial B pH<2	JC/MD	Ok
13	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045138.D	05 Mar 2025 13:36	Not req.	JC/MD	Not Ok
14	Q1478-08	IDW-AQ-DRUM-612-02	VX045139.D	05 Mar 2025 13:59	Need lower dilution	JC/MD	Not Ok
15	VX0305MBSD01	VX0305MBSD01	VX045140.D	05 Mar 2025 14:23		JC/MD	Ok,M
16	Q1448-04MEDL	P3MEDL	VX045141.D	05 Mar 2025 14:46	Not req.	JC/MD	Not Ok
17	IBLK	IBLK	VX045142.D	05 Mar 2025 15:09		JC/MD	Ok
18	Q1478-08	IDW-AQ-DRUM-612-02	VX045143.D	05 Mar 2025 15:32	Need lower dilution	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030525

Review By	John Carlone	Review On	3/6/2025 8:28:41 AM
Supervise By	Maresh Dadoda	Supervise On	3/6/2025 2:56:36 PM
SubDirectory	VX030525	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133217		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133218,VP133219		

19	Q1478-02DL	IDW-AQ-DRUM-610-02	VX045144.D	05 Mar 2025 15:56	vial B pH<2	JC/MD	Ok
20	Q1478-08	IDW-AQ-DRUM-612-02	VX045145.D	05 Mar 2025 16:19	vial A pH<2	JC/MD	Ok
21	Q1501-03	EB01-030525	VX045146.D	05 Mar 2025 16:42	vial A pH<2 EB	JC/MD	Ok
22	Q1501-07	TB01-030525	VX045147.D	05 Mar 2025 17:05	vial A pH<2 TB	JC/MD	Ok
23	Q1501-04	BR-05-465-030525	VX045148.D	05 Mar 2025 17:28	Need Straight Run	JC/MD	Not Ok
24	Q1488-15	TB03042025	VX045149.D	05 Mar 2025 17:52	vial A pH<2 TB	JC/MD	Ok
25	Q1488-13	ENV-102-GW01	VX045150.D	05 Mar 2025 18:15	vial A pH<2	JC/MD	Ok
26	Q1488-14	ENV-104-GW01	VX045151.D	05 Mar 2025 18:38	vial A pH<2 Need 4X	JC/MD	Dilution
27	Q1494-01	PURGE-WATER	VX045152.D	05 Mar 2025 19:01	vial A pH<2 E flag in previous sample	JC/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VX045153.D	05 Mar 2025 19:24		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030625

Review By	John Carlone	Review On	3/10/2025 9:58:48 AM
Supervise By	Maresh Dadoda	Supervise On	3/10/2025 2:27:03 PM
SubDirectory	VX030625	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133226
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133227,VP133228

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045154.D	06 Mar 2025 09:09		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045155.D	06 Mar 2025 09:56	pH#Lot#V12668	JC/MD	Ok,M
3	VX0306MBL01	VX0306MBL01	VX045156.D	06 Mar 2025 10:23		JC/MD	Ok
4	VX0306WBL01	VX0306WBL01	VX045157.D	06 Mar 2025 10:46		JC/MD	Ok
5	VX0306WBS01	VX0306WBS01	VX045158.D	06 Mar 2025 11:10		JC/MD	Ok,M
6	Q1501-01	MW-16B-87.5-030425	VX045159.D	06 Mar 2025 11:36	vial A pH<2	JC/MD	Ok
7	Q1501-02	MW-16B-87.5-030425	VX045160.D	06 Mar 2025 12:00	vial A pH<2	JC/MD	Ok
8	Q1488-14DL	ENV-104-GW01DL	VX045161.D	06 Mar 2025 12:23	vial B pH<2	JC/MD	Ok,M
9	Q1501-04	BR-05-465-030525	VX045162.D	06 Mar 2025 12:46	vial A pH<2 Surrogate Fail	JC/MD	ReRun
10	Q1501-05MS	BR-05-465-030525MS	VX045163.D	06 Mar 2025 13:09	vial A pH<2	JC/MD	Ok,M
11	Q1501-06MSD	BR-05-465-030525MSD	VX045164.D	06 Mar 2025 13:33	vial A pH<2	JC/MD	Ok,M
12	Q1484-06	TR-OTR-COMP-01	VX045165.D	06 Mar 2025 13:56	vial A pH#5.0	JC/MD	Ok
13	Q1494-01	PURGE-WATER	VX045166.D	06 Mar 2025 14:19	vial B pH<2	JC/MD	Ok
14	Q1501-04RE	BR-05-465-030525RE	VX045167.D	06 Mar 2025 14:43	vial B pH<2 Surrogate Fail	JC/MD	Confirms
15	PB166989TB	PB166989TB	VX045168.D	06 Mar 2025 15:06		JC/MD	Ok
16	Q1481-02	WC-K1311	VX045169.D	06 Mar 2025 15:29	vial A pH#5.0	JC/MD	Ok,M
17	Q1488-02	ENV-101-SB01	VX045170.D	06 Mar 2025 15:52	vial A pH#5.0	JC/MD	Ok
18	Q1488-04	ENV-101-SB02	VX045171.D	06 Mar 2025 16:16	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX030625

Review By	John Carlone	Review On	3/10/2025 9:58:48 AM
Supervise By	Maresh Dadoda	Supervise On	3/10/2025 2:27:03 PM
SubDirectory	VX030625	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133226		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133227,VP133228		

19	Q1488-06	ENV-102-SB01	VX045172.D	06 Mar 2025 16:39	vial A pH#5.0	JC/MD	Ok
20	Q1488-08	ENV-102-SB02	VX045173.D	06 Mar 2025 17:02	vial A pH#5.0	JC/MD	Ok
21	Q1488-10	ENV-104-SB01	VX045174.D	06 Mar 2025 17:25	vial A pH#5.0	JC/MD	Ok
22	Q1488-12	ENV-104-SB02	VX045175.D	06 Mar 2025 17:49	vial A pH#5.0	JC/MD	Ok
23	Q1490-02	CLAY-SLUDGE-DRUM	VX045176.D	06 Mar 2025 18:12	vial A pH#5.0	JC/MD	Ok
24	Q1494-04	ASPHALT-SOIL	VX045177.D	06 Mar 2025 18:35	vial A pH#5.0	JC/MD	Ok
25	Q1494-06	SOIL	VX045178.D	06 Mar 2025 18:59	vial A pH#5.0	JC/MD	Ok
26	Q1494-08	SOIL-COMP	VX045179.D	06 Mar 2025 19:22	vial A pH#5.0	JC/MD	Ok
27	Q1478-15	IDW-SO-DRUM-582-02	VX045180.D	06 Mar 2025 19:45	vial A pH#5.0	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX045181.D	06 Mar 2025 20:09		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Mahesh Dadoda	Review On	3/5/2025 12:14:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM		
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133216
Initial Calibration Stds	VP133207,VP133208,VP133209,VP133210,VP133211,VP133212
CCC	VP133214,VP133215
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133213
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133223
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133224,VP133225 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021425.D	06 Mar 2025 09:39		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021426.D	06 Mar 2025 10:10		SY/MD	Ok,M
3	VY0306SBL01	VY0306SBL01	VY021427.D	06 Mar 2025 11:59		SY/MD	Ok
4	VY0306SBS01	VY0306SBS01	VY021428.D	06 Mar 2025 12:33		SY/MD	Ok,M
5	VY0306SBSD01	VY0306SBSD01	VY021429.D	06 Mar 2025 12:55		SY/MD	Ok,M
6	Q1489-05	TP-2	VY021430.D	06 Mar 2025 13:28	vial-B	SY/MD	Ok
7	Q1489-01	TP-1	VY021431.D	06 Mar 2025 13:52	vial-B Not purge	SY/MD	Not Ok
8	Q1421-09	P001-CLAY-CF02-01	VY021432.D	06 Mar 2025 14:15	vial-B	SY/MD	Ok
9	Q1479-01	P5	VY021433.D	06 Mar 2025 14:39	vial-A	SY/MD	Ok
10	Q1479-02	DSP2	VY021434.D	06 Mar 2025 15:02	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q1488-01	ENV-101-SB01	VY021435.D	06 Mar 2025 15:25	vial-A	SY/MD	Ok
12	Q1488-03	ENV-101-SB02	VY021436.D	06 Mar 2025 15:49	vial-A	SY/MD	Ok
13	Q1488-05	ENV-102-SB01	VY021437.D	06 Mar 2025 16:12	vial-A	SY/MD	Ok
14	Q1488-07	ENV-102-SB02	VY021438.D	06 Mar 2025 16:36	vial-A Not purge	SY/MD	Not Ok
15	Q1488-09	ENV-104-SB01	VY021439.D	06 Mar 2025 16:59	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q1488-11	ENV-104-SB02	VY021440.D	06 Mar 2025 17:22	vial-A Not purge	SY/MD	Not Ok
17	Q1463-01	T1	VY021441.D	06 Mar 2025 17:46	vial-B	SY/MD	Ok
18	Q1463-05	T5	VY021442.D	06 Mar 2025 18:09	vial-B	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030625

Review By	Maresh Dadoda	Review On	3/7/2025 1:32:16 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/7/2025 1:32:53 PM
SubDirectory	VY030625	HP Acquire Method	HP Processing Method 82y030425s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133223		
CCC	VP133224,VP133225		
Internal Standard/PEM	VP131783		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1463-04	T4	VY021443.D	06 Mar 2025 18:33	Vial-B	SY/MD	Ok
20	Q1463-06	T6	VY021444.D	06 Mar 2025 18:56	Vial-A	SY/MD	Ok
21	Q1463-07	T7	VY021445.D	06 Mar 2025 19:20	Vial-A	SY/MD	Ok
22	Q1463-08	T8	VY021446.D	06 Mar 2025 19:43	Vial-A	SY/MD	Ok
23	Q1463-09	T9	VY021447.D	06 Mar 2025 20:06	Vial-A	SY/MD	Ok
24	Q1463-10	T10	VY021448.D	06 Mar 2025 20:30	Vial-A Internal Standard Fail; Surrogate Fail	SY/MD	ReRun
25	VSTDCCC050	VSTDCCC050EC	VY021449.D	06 Mar 2025 21:15		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030725

Review By	Mahesh Dadoda	Review On	3/11/2025 2:51:48 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/11/2025 2:52:31 PM		
SubDirectory	VY030725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133232
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133233,VP133234 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021450.D	07 Mar 2025 09:27		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021451.D	07 Mar 2025 09:59		SY/MD	Ok,M
3	VY0307SBL01	VY0307SBL01	VY021452.D	07 Mar 2025 10:35		SY/MD	Ok
4	VY0307SBS01	VY0307SBS01	VY021453.D	07 Mar 2025 12:03		SY/MD	Ok,M
5	VY0307SBSD01	VY0307SBSD01	VY021454.D	07 Mar 2025 12:26	BSD Failed Low for com.#73,87,88	SY/MD	Ok,M
6	Q1463-03	T3	VY021455.D	07 Mar 2025 12:55	vial-B	SY/MD	Ok,M
7	Q1479-02	DSP2	VY021456.D	07 Mar 2025 13:18	vial-B	SY/MD	Ok,M
8	Q1463-02	T2	VY021457.D	07 Mar 2025 13:42	vial-A Internal standard fail	SY/MD	ReRun
9	Q1463-10	T10	VY021458.D	07 Mar 2025 14:05	vial-B	SY/MD	Ok
10	Q1488-07	ENV-102-SB02	VY021459.D	07 Mar 2025 14:28	vial-B	SY/MD	Ok,M
11	Q1488-09	ENV-104-SB01	VY021460.D	07 Mar 2025 14:52	Vial-B	SY/MD	Ok
12	Q1488-11	ENV-104-SB02	VY021461.D	07 Mar 2025 15:15	Vial-B Internal standard fail	SY/MD	Ok
13	Q1514-01	ENV-105-SB01	VY021462.D	07 Mar 2025 15:39	Vial-A BSD failed Low	SY/MD	ReRun
14	Q1514-03	ENV-105-SB02	VY021463.D	07 Mar 2025 16:02	Vial-A BSD failed Low	SY/MD	ReRun
15	Q1514-05	ENV-103-SB01	VY021464.D	07 Mar 2025 16:25	Vial-A BSD failed Low	SY/MD	ReRun
16	Q1515-02	AU-06-030625-E2	VY021465.D	07 Mar 2025 16:49	Vial-A Internal standard fail;Surrogate fail	SY/MD	ReRun
17	Q1507-01	50-MIDDLESEX-AVE	VY021466.D	07 Mar 2025 17:12	Vial-A Not purge	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030725

Review By	Mahesh Dadoda	Review On	3/11/2025 2:51:48 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/11/2025 2:52:31 PM		
SubDirectory	VY030725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y030425s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133232			
CCC		VP133233,VP133234			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1508-01	RBR251372	VY021467.D	07 Mar 2025 17:36	Vial-A Internal standard fail	SY/MD	ReRun
19	Q1494-05	SOIL	VY021468.D	07 Mar 2025 17:59	Vial-A Internal standard fail	SY/MD	ReRun
20	Q1494-03	ASPHALT-SOIL	VY021469.D	07 Mar 2025 18:23	Vial-A Internal standard fail	SY/MD	ReRun
21	Q1494-07	SOIL-COMP	VY021470.D	07 Mar 2025 18:46	Vial-A Internal standard fail	SY/MD	ReRun
22	Q1490-01	CLAY-SLUDGE-DRUM	VY021471.D	07 Mar 2025 19:09	Vial-A Not purge	SY/MD	Not Ok
23	VSTDCCC050	VSTDCCC050EC	VY021472.D	07 Mar 2025 19:32		SY/MD	Ok,M

M : Manual Integration