

Surrogate Summary

SDG No.: **Q1488**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1488-01	ENV-101-SB01	1,2-Dichloroethane-d4	50	65.1	130	70 (63)	130 (155)
		Dibromofluoromethane	50	52.5	105	70 (70)	130 (134)
		Toluene-d8	50	48.8	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	45.9	92	70 (38)	130 (136)
Q1488-03	ENV-101-SB02	1,2-Dichloroethane-d4	50	60.7	121	70 (63)	130 (155)
		Dibromofluoromethane	50	51.7	103	70 (70)	130 (134)
		Toluene-d8	50	48.8	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.8	86	70 (38)	130 (136)
Q1488-05	ENV-102-SB01	1,2-Dichloroethane-d4	50	64.0	128	70 (63)	130 (155)
		Dibromofluoromethane	50	52.7	105	70 (70)	130 (134)
		Toluene-d8	50	49.2	98	70 (74)	130 (123)
		4-Bromofluorobenzene	50	46.3	93	70 (38)	130 (136)
Q1488-07	ENV-102-SB02	1,2-Dichloroethane-d4	50	56.9	114	70 (63)	130 (155)
		Dibromofluoromethane	50	50.5	101	70 (70)	130 (134)
		Toluene-d8	50	47.9	96	70 (74)	130 (123)
		4-Bromofluorobenzene	50	42.0	84	70 (38)	130 (136)
Q1488-09	ENV-104-SB01	1,2-Dichloroethane-d4	50	55.9	112	70 (63)	130 (155)
		Dibromofluoromethane	50	52.3	105	70 (70)	130 (134)
		Toluene-d8	50	48.5	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	33.0	66 *	70 (38)	130 (136)
Q1488-11	ENV-104-SB02	1,2-Dichloroethane-d4	50	55.2	110	70 (63)	130 (155)
		Dibromofluoromethane	50	52.2	104	70 (70)	130 (134)
		Toluene-d8	50	44.3	89	70 (74)	130 (123)
		4-Bromofluorobenzene	50	26.3	53 *	70 (38)	130 (136)
VY0306SBL01	VY0306SBL01	1,2-Dichloroethane-d4	50	58.5	117	70 (63)	130 (155)
		Dibromofluoromethane	50	51.1	102	70 (70)	130 (134)
		Toluene-d8	50	48.6	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	43.6	87	70 (38)	130 (136)
VY0306SBS01	VY0306SBS01	1,2-Dichloroethane-d4	50	53.2	106	70 (63)	130 (155)
		Dibromofluoromethane	50	50.1	100	70 (70)	130 (134)
		Toluene-d8	50	49.5	99	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.8	102	70 (38)	130 (136)
VY0306SBSD01	VY0306SBSD01	1,2-Dichloroethane-d4	50	55.0	110	70 (63)	130 (155)
		Dibromofluoromethane	50	52.0	104	70 (70)	130 (134)
		Toluene-d8	50	51.4	103	70 (74)	130 (123)
		4-Bromofluorobenzene	50	52.6	105	70 (38)	130 (136)
VY0307SBL01	VY0307SBL01	1,2-Dichloroethane-d4	50	55.4	111	70 (63)	130 (155)
		Dibromofluoromethane	50	50.2	100	70 (70)	130 (134)
		Toluene-d8	50	48.5	97	70 (74)	130 (123)
		4-Bromofluorobenzene	50	40.1	80	70 (38)	130 (136)
VY0307SBS01	VY0307SBS01	1,2-Dichloroethane-d4	50	53.1	106	70 (63)	130 (155)
		Dibromofluoromethane	50	52.0	104	70 (70)	130 (134)
		Toluene-d8	50	51.9	104	70 (74)	130 (123)
		4-Bromofluorobenzene	50	50.9	102	70 (38)	130 (136)

Surrogate Summary

SDG No.: Q1488

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1488-13	ENV-102-GW01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.3	109	70 (77)	130 (121)
Q1488-14	ENV-104-GW01	1,2-Dichloroethane-d4	50	51.8	104	70 (74)	130 (125)
		Dibromofluoromethane	50	51.1	102	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
Q1488-14DL	ENV-104-GW01DL	1,2-Dichloroethane-d4	50	52.0	104	70 (74)	130 (125)
		Dibromofluoromethane	50	53.1	106	70 (75)	130 (124)
		Toluene-d8	50	53.4	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.8	104	70 (77)	130 (121)
Q1488-15	TB03042025	1,2-Dichloroethane-d4	50	53.4	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VX0305WBL01	VX0305WBL01	1,2-Dichloroethane-d4	50	52.7	105	70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)
VX0305WBS01	VX0305WBS01	1,2-Dichloroethane-d4	50	48.4	97	70 (74)	130 (125)
		Dibromofluoromethane	50	49.3	99	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.5	99	70 (77)	130 (121)
VX0305WBSD0	VX0305WBSD01	1,2-Dichloroethane-d4	50	49.3	99	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.2	102	70 (77)	130 (121)
VX0306WBL01	VX0306WBL01	1,2-Dichloroethane-d4	50	53.9	108	70 (74)	130 (125)
		Dibromofluoromethane	50	52.4	105	70 (75)	130 (124)
		Toluene-d8	50	52.3	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.8	104	70 (77)	130 (121)
VX0306WBS01	VX0306WBS01	1,2-Dichloroethane-d4	50	50.2	100	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.2	98	70 (77)	130 (121)

