

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488
 Lab File ID: VX045155.D Date Analyzed: 03/06/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:56
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	99532	5.54	175834	6.76	153746	10.05
UPPER LIMIT	199064	6.044	351668	7.257	307492	10.549
LOWER LIMIT	49766	5.044	87917	6.257	76873	9.549
EPA SAMPLE NO.						
ENV-104-GW01DL	70228	5.55	133297	6.76	124634	10.05
VX0306WBL01	71177	5.54	139793	6.76	129041	10.05
VX0306WBS01	97102	5.54	175752	6.76	147757	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.