

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488
 Lab File ID: VY021426.D Date Analyzed: 03/06/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	228274	7.71	348529	8.61	321426	11.41
UPPER LIMIT	456548	8.207	697058	9.11	642852	11.914
LOWER LIMIT	114137	7.207	174265	8.11	160713	10.914
EPA SAMPLE NO.						
ENV-101-SB01	242903	7.71	455891	8.62	410551	11.41
ENV-101-SB02	234226	7.71	429130	8.62	377679	11.41
ENV-102-SB01	245112	7.71	450789	8.62	408673	11.41
VY0306SBL01	249443	7.71	460921	8.62	405928	11.42
VY0306SBS01	200172	7.71	318431	8.62	292608	11.42
VY0306SBSD01	197190	7.71	317024	8.62	284896	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.