

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488
 Lab File ID: VY021451.D Date Analyzed: 03/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	212653	7.71	324643	8.62	291936	11.42
UPPER LIMIT	425306	8.207	649286	9.116	583872	11.92
LOWER LIMIT	106327	7.207	162322	8.116	145968	10.92
EPA SAMPLE NO.						
ENV-102-SB02	280904	7.71	508130	8.62	440256	11.41
ENV-104-SB01	272187	7.71	479219	8.62	375437	11.41
ENV-104-SB02	263229	7.71	450669	8.62	301719	11.41
VY0307SBL01	247432	7.71	455944	8.62	386511	11.42
VY0307SBS01	219868	7.71	343563	8.62	301562	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.