

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488
 Lab File ID: VX045155.D Date Analyzed: 03/06/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:56
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	69403	12.018				
UPPER LIMIT	138806	12.518				
LOWER LIMIT	34701.5	11.518				
EPA SAMPLE NO.						
ENV-104-GW01DL	48192	12.02				
VX0306WBL01	53719	12.02				
VX0306WBS01	67123	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.