

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: Q1488 SAS No.: Q1488 SDG NO.: Q1488  
 Lab File ID: VY021451.D Date Analyzed: 03/07/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 09:59  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 151541        | 13.353 |  |  |  |  |
| UPPER LIMIT    | 303082        | 13.853 |  |  |  |  |
| LOWER LIMIT    | 75770.5       | 12.853 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| ENV-102-SB02   | 168898        | 13.35  |  |  |  |  |
| ENV-104-SB01   | 106159        | 13.35  |  |  |  |  |
| ENV-104-SB02   | 63186 *       | 13.35  |  |  |  |  |
| VY0307SBL01    | 140975        | 13.35  |  |  |  |  |
| VY0307SBS01    | 152803        | 13.35  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.