

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

# **CASE NARRATIVE**

**Aramark Uniforms** 

**Project Name: Monthly 2025** 

Project # N/A

**Chemtech Project # Q1491** 

Test Name: Metals ICP-Group1, Mercury

### A. Number of Samples and Date of Receipt:

2 Water samples were received on 03/05/2025.

#### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: BOD5, Mercury, Metals Group1, Metals ICP-Group1, TPH and TSS. This data package contains results for Metals ICP-Group1, Mercury.

## C. Analytical Techniques:

The analysis and digestion of Metals ICP-Group1 was based on 200.7 and The analysis and digestion of Mercury was based on 245.1.

## D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met criteria for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

#### **E. Additional Comments:**

As per special requirement for this project form-1 and Hit summary are reported in mg/l.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.