

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1491

Project: Monthly 2025

Run No.: LB134910

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134910BS							
BOD5	mg/L	198	185		93	1	84.6-115.4	03/05/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1491

Project: Monthly 2025

Run No.: LB134923

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134923BS							
TSS	mg/L	550	533		97	1	90-110	03/06/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1491

Project: Monthly 2025

Run No.: LB134936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134936BS							
TPH	mg/L	20.0	16.9		84	1	78-114	03/07/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1491

Project: Monthly 2025

Run No.: LB134936

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134936BSD							
TPH	mg/L	20.0	17.1		86	1	78-114	03/07/2025