

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV41	Mercury	4.17	4.0	104	95 - 105	CV	03/13/2025	11:20	LB135019

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Mercury	5.32	5.0	106	90 - 110	CV	03/13/2025	11:24	LB135019
CCV07	Mercury	5.25	5.0	105	90 - 110	CV	03/13/2025	12:10	LB135019
CCV08	Mercury	5.02	5.0	100	90 - 110	CV	03/13/2025	12:22	LB135019

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	511	510	100	95 - 105	P	03/11/2025	10:50	LB134987
	Copper	532	510	104	95 - 105	P	03/11/2025	10:50	LB134987
	Lead	1020	1000	102	95 - 105	P	03/11/2025	10:50	LB134987
	Nickel	517	530	98	95 - 105	P	03/11/2025	10:50	LB134987
	Zinc	1040	1000	104	95 - 105	P	03/11/2025	10:50	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.80	6.0	97	80 - 120	P	03/11/2025	10:54	LB134987
	Copper	22.1	20.0	110	80 - 120	P	03/11/2025	10:54	LB134987
	Lead	12.0	12.0	100	80 - 120	P	03/11/2025	10:54	LB134987
	Nickel	38.4	40.0	96	80 - 120	P	03/11/2025	10:54	LB134987
	Zinc	33.7	40.0	84	80 - 120	P	03/11/2025	10:54	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2520	2500	101	90 - 110	P	03/11/2025	11:29	LB134987
	Copper	1280	1250	102	90 - 110	P	03/11/2025	11:29	LB134987
	Lead	5030	5000	101	90 - 110	P	03/11/2025	11:29	LB134987
	Nickel	2510	2500	100	90 - 110	P	03/11/2025	11:29	LB134987
	Zinc	2560	2500	102	90 - 110	P	03/11/2025	11:29	LB134987
CCV02	Cadmium	2570	2500	103	90 - 110	P	03/11/2025	12:19	LB134987
	Copper	1300	1250	104	90 - 110	P	03/11/2025	12:19	LB134987
	Lead	5120	5000	102	90 - 110	P	03/11/2025	12:19	LB134987
	Nickel	2550	2500	102	90 - 110	P	03/11/2025	12:19	LB134987
	Zinc	2620	2500	105	90 - 110	P	03/11/2025	12:19	LB134987
CCV03	Cadmium	2430	2500	97	90 - 110	P	03/11/2025	13:10	LB134987
	Copper	1250	1250	100	90 - 110	P	03/11/2025	13:10	LB134987
	Lead	4830	5000	97	90 - 110	P	03/11/2025	13:10	LB134987
	Nickel	2410	2500	96	90 - 110	P	03/11/2025	13:10	LB134987
	Zinc	2480	2500	99	90 - 110	P	03/11/2025	13:10	LB134987
CCV04	Cadmium	2470	2500	99	90 - 110	P	03/11/2025	14:08	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	14:08	LB134987
	Lead	4880	5000	98	90 - 110	P	03/11/2025	14:08	LB134987
	Nickel	2430	2500	97	90 - 110	P	03/11/2025	14:08	LB134987
	Zinc	2530	2500	101	90 - 110	P	03/11/2025	14:08	LB134987
CCV05	Cadmium	2400	2500	96	90 - 110	P	03/11/2025	15:00	LB134987
	Copper	1240	1250	99	90 - 110	P	03/11/2025	15:00	LB134987
	Lead	4760	5000	95	90 - 110	P	03/11/2025	15:00	LB134987
	Nickel	2360	2500	94	90 - 110	P	03/11/2025	15:00	LB134987
	Zinc	2430	2500	97	90 - 110	P	03/11/2025	15:00	LB134987
CCV06	Cadmium	2430	2500	97	90 - 110	P	03/11/2025	15:50	LB134987
	Copper	1240	1250	99	90 - 110	P	03/11/2025	15:50	LB134987
	Lead	4800	5000	96	90 - 110	P	03/11/2025	15:50	LB134987
	Nickel	2390	2500	95	90 - 110	P	03/11/2025	15:50	LB134987
	Zinc	2490	2500	100	90 - 110	P	03/11/2025	15:50	LB134987
CCV07	Cadmium	2460	2500	98	90 - 110	P	03/11/2025	16:42	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	16:42	LB134987
	Lead	4850	5000	97	90 - 110	P	03/11/2025	16:42	LB134987
	Nickel	2400	2500	96	90 - 110	P	03/11/2025	16:42	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2490	2500	99	90 - 110	P	03/11/2025	16:42	LB134987
CCV08	Cadmium	2480	2500	99	90 - 110	P	03/11/2025	17:34	LB134987
	Copper	1270	1250	102	90 - 110	P	03/11/2025	17:34	LB134987
	Lead	4880	5000	98	90 - 110	P	03/11/2025	17:34	LB134987
	Nickel	2420	2500	97	90 - 110	P	03/11/2025	17:34	LB134987
	Zinc	2500	2500	100	90 - 110	P	03/11/2025	17:34	LB134987
CCV09	Cadmium	2520	2500	101	90 - 110	P	03/11/2025	18:26	LB134987
	Copper	1290	1250	103	90 - 110	P	03/11/2025	18:26	LB134987
	Lead	4960	5000	99	90 - 110	P	03/11/2025	18:26	LB134987
	Nickel	2460	2500	98	90 - 110	P	03/11/2025	18:26	LB134987
	Zinc	2550	2500	102	90 - 110	P	03/11/2025	18:26	LB134987
CCV10	Cadmium	2480	2500	99	90 - 110	P	03/11/2025	19:14	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	19:14	LB134987
	Lead	4880	5000	98	90 - 110	P	03/11/2025	19:14	LB134987
	Nickel	2420	2500	97	90 - 110	P	03/11/2025	19:14	LB134987
	Zinc	2480	2500	99	90 - 110	P	03/11/2025	19:14	LB134987
CCV11	Cadmium	2390	2500	95	90 - 110	P	03/11/2025	19:26	LB134987
	Copper	1240	1250	100	90 - 110	P	03/11/2025	19:26	LB134987
	Lead	4740	5000	95	90 - 110	P	03/11/2025	19:26	LB134987
	Nickel	2350	2500	94	90 - 110	P	03/11/2025	19:26	LB134987
	Zinc	2470	2500	99	90 - 110	P	03/11/2025	19:26	LB134987
CCV12	Cadmium	2380	2500	95	90 - 110	P	03/11/2025	19:39	LB134987
	Copper	1230	1250	99	90 - 110	P	03/11/2025	19:39	LB134987
	Lead	4680	5000	94	90 - 110	P	03/11/2025	19:39	LB134987
	Nickel	2350	2500	94	90 - 110	P	03/11/2025	19:39	LB134987
	Zinc	2430	2500	97	90 - 110	P	03/11/2025	19:39	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	503	510	99	95 - 105	P	03/13/2025	12:07	LB135035
	Copper	529	510	104	95 - 105	P	03/13/2025	12:07	LB135035
	Lead	1000	1000	100	95 - 105	P	03/13/2025	12:07	LB135035
	Nickel	545	530	103	95 - 105	P	03/13/2025	12:07	LB135035
	Zinc	1020	1000	102	95 - 105	P	03/13/2025	12:07	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.95	6.0	99	80 - 120	P	03/13/2025	12:13	LB135035
	Copper	22.7	20.0	114	80 - 120	P	03/13/2025	12:13	LB135035
	Lead	12.6	12.0	105	80 - 120	P	03/13/2025	12:13	LB135035
	Nickel	39.7	40.0	99	80 - 120	P	03/13/2025	12:13	LB135035
	Zinc	41.3	40.0	103	80 - 120	P	03/13/2025	12:13	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	12:53	LB135035
	Copper	1280	1250	103	90 - 110	P	03/13/2025	12:53	LB135035
	Lead	5190	5000	104	90 - 110	P	03/13/2025	12:53	LB135035
	Nickel	2560	2500	102	90 - 110	P	03/13/2025	12:53	LB135035
	Zinc	2600	2500	104	90 - 110	P	03/13/2025	12:53	LB135035
CCV02	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	13:49	LB135035
	Copper	1310	1250	105	90 - 110	P	03/13/2025	13:49	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	13:49	LB135035
	Nickel	2600	2500	104	90 - 110	P	03/13/2025	13:49	LB135035
	Zinc	2620	2500	105	90 - 110	P	03/13/2025	13:49	LB135035
CCV03	Cadmium	2540	2500	102	90 - 110	P	03/13/2025	14:40	LB135035
	Copper	1270	1250	102	90 - 110	P	03/13/2025	14:40	LB135035
	Lead	5150	5000	103	90 - 110	P	03/13/2025	14:40	LB135035
	Nickel	2530	2500	101	90 - 110	P	03/13/2025	14:40	LB135035
	Zinc	2560	2500	102	90 - 110	P	03/13/2025	14:40	LB135035
CCV04	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	15:38	LB135035
	Copper	1240	1250	99	90 - 110	P	03/13/2025	15:38	LB135035
	Lead	5050	5000	101	90 - 110	P	03/13/2025	15:38	LB135035
	Nickel	2480	2500	99	90 - 110	P	03/13/2025	15:38	LB135035
	Zinc	2460	2500	98	90 - 110	P	03/13/2025	15:38	LB135035
CCV05	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	16:28	LB135035
	Copper	1300	1250	104	90 - 110	P	03/13/2025	16:28	LB135035
	Lead	5300	5000	106	90 - 110	P	03/13/2025	16:28	LB135035
	Nickel	2620	2500	105	90 - 110	P	03/13/2025	16:28	LB135035
	Zinc	2580	2500	103	90 - 110	P	03/13/2025	16:28	LB135035
CCV06	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	17:21	LB135035
	Copper	1270	1250	102	90 - 110	P	03/13/2025	17:21	LB135035
	Lead	5180	5000	104	90 - 110	P	03/13/2025	17:21	LB135035
	Nickel	2550	2500	102	90 - 110	P	03/13/2025	17:21	LB135035
	Zinc	2550	2500	102	90 - 110	P	03/13/2025	17:21	LB135035
CCV07	Cadmium	2640	2500	106	90 - 110	P	03/13/2025	18:12	LB135035
	Copper	1310	1250	105	90 - 110	P	03/13/2025	18:12	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	18:12	LB135035
	Nickel	2620	2500	105	90 - 110	P	03/13/2025	18:12	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2600	2500	104	90 - 110	P	03/13/2025	18:12	LB135035
CCV08	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	18:53	LB135035
	Copper	1240	1250	99	90 - 110	P	03/13/2025	18:53	LB135035
	Lead	5060	5000	101	90 - 110	P	03/13/2025	18:53	LB135035
	Nickel	2480	2500	99	90 - 110	P	03/13/2025	18:53	LB135035
	Zinc	2330	2500	93	90 - 110	P	03/13/2025	18:53	LB135035
CCV09	Cadmium	2550	2500	102	90 - 110	P	03/13/2025	19:12	LB135035
	Copper	1290	1250	103	90 - 110	P	03/13/2025	19:12	LB135035
	Lead	5120	5000	102	90 - 110	P	03/13/2025	19:12	LB135035
	Nickel	2550	2500	102	90 - 110	P	03/13/2025	19:12	LB135035
	Zinc	2490	2500	100	90 - 110	P	03/13/2025	19:12	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	504	510	99	95 - 105	P	03/17/2025	14:31	LB135063
	Copper	528	510	104	95 - 105	P	03/17/2025	14:31	LB135063
	Lead	996	1000	100	95 - 105	P	03/17/2025	14:31	LB135063
	Nickel	512	530	97	95 - 105	P	03/17/2025	14:31	LB135063
	Zinc	1040	1000	104	95 - 105	P	03/17/2025	14:31	LB135063

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	5.89	6.0	98	80 - 120	P	03/17/2025	14:38	LB135063
	Copper	22.1	20.0	110	80 - 120	P	03/17/2025	14:38	LB135063
	Lead	12.3	12.0	103	80 - 120	P	03/17/2025	14:38	LB135063
	Nickel	39.3	40.0	98	80 - 120	P	03/17/2025	14:38	LB135063
	Zinc	42.5	40.0	106	80 - 120	P	03/17/2025	14:38	LB135063

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2510	2500	100	90 - 110	P	03/17/2025	15:15	LB135063
	Copper	1270	1250	102	90 - 110	P	03/17/2025	15:15	LB135063
	Lead	5080	5000	102	90 - 110	P	03/17/2025	15:15	LB135063
	Nickel	2490	2500	100	90 - 110	P	03/17/2025	15:15	LB135063
	Zinc	2620	2500	105	90 - 110	P	03/17/2025	15:15	LB135063
CCV02	Cadmium	2500	2500	100	90 - 110	P	03/17/2025	16:13	LB135063
	Copper	1280	1250	102	90 - 110	P	03/17/2025	16:13	LB135063
	Lead	5070	5000	101	90 - 110	P	03/17/2025	16:13	LB135063
	Nickel	2470	2500	99	90 - 110	P	03/17/2025	16:13	LB135063
	Zinc	2590	2500	104	90 - 110	P	03/17/2025	16:13	LB135063
CCV03	Cadmium	2570	2500	103	90 - 110	P	03/17/2025	17:05	LB135063
	Copper	1310	1250	104	90 - 110	P	03/17/2025	17:05	LB135063
	Lead	5200	5000	104	90 - 110	P	03/17/2025	17:05	LB135063
	Nickel	2530	2500	101	90 - 110	P	03/17/2025	17:05	LB135063
	Zinc	2680	2500	107	90 - 110	P	03/17/2025	17:05	LB135063
CCV04	Cadmium	2390	2500	96	90 - 110	P	03/17/2025	17:51	LB135063
	Copper	1230	1250	98	90 - 110	P	03/17/2025	17:51	LB135063
	Lead	4870	5000	97	90 - 110	P	03/17/2025	17:51	LB135063
	Nickel	2360	2500	94	90 - 110	P	03/17/2025	17:51	LB135063
	Zinc	2510	2500	100	90 - 110	P	03/17/2025	17:51	LB135063
CCV05	Cadmium	2390	2500	96	90 - 110	P	03/17/2025	18:18	LB135063
	Copper	1240	1250	99	90 - 110	P	03/17/2025	18:18	LB135063
	Lead	4870	5000	97	90 - 110	P	03/17/2025	18:18	LB135063
	Nickel	2360	2500	94	90 - 110	P	03/17/2025	18:18	LB135063
	Zinc	2540	2500	102	90 - 110	P	03/17/2025	18:18	LB135063