



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms SDG No.: Q1491
Contract: ARAM01 Lab Code: CHEM Case No.: Q1491 SAS No.: Q1491
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.81	6.0	97	40 - 160	P	03/11/2025	11:02	LB134987
	Copper	22.1	20.0	110	40 - 160	P	03/11/2025	11:02	LB134987
	Lead	11.9	12.0	99	40 - 160	P	03/11/2025	11:02	LB134987
	Nickel	38.2	40.0	96	40 - 160	P	03/11/2025	11:02	LB134987
	Zinc	33.2	40.0	83	40 - 160	P	03/11/2025	11:02	LB134987
CRA	Mercury	0.059	0.05	119	40 - 160	CV	03/13/2025	11:40	LB135019
CRI01	Cadmium	5.90	6.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Copper	23.1	20.0	116	40 - 160	P	03/13/2025	12:21	LB135035
	Lead	11.7	12.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Nickel	39.8	40.0	100	40 - 160	P	03/13/2025	12:21	LB135035
	Zinc	42.5	40.0	106	40 - 160	P	03/13/2025	12:21	LB135035
CRI01	Cadmium	5.88	6.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Copper	22.2	20.0	111	40 - 160	P	03/17/2025	14:47	LB135063
	Lead	12.6	12.0	105	40 - 160	P	03/17/2025	14:47	LB135063
	Nickel	39.4	40.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Zinc	42.9	40.0	107	40 - 160	P	03/17/2025	14:47	LB135063