

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-2.31	1.0	231	-5	7	03/11/2025	11:06	LB134987
	Copper	6.76	2.0	338	-18	22	03/11/2025	11:06	LB134987
	Lead	3.81			-12	12	03/11/2025	11:06	LB134987
	Nickel	1.76	2.0	88	-38	42	03/11/2025	11:06	LB134987
	Zinc	-3.79			-40	40	03/11/2025	11:06	LB134987
ICSAB01	Cadmium	1020	972	105	826	1120	03/11/2025	11:11	LB134987
	Copper	507	511	99	434	588	03/11/2025	11:11	LB134987
	Lead	55.7	49.0	114	37	61	03/11/2025	11:11	LB134987
	Nickel	1020	954	107	810	1100	03/11/2025	11:11	LB134987
	Zinc	1070	952	112	809	1095	03/11/2025	11:11	LB134987
ICSA	Cadmium	-2.80	1.0	280	-5	7	03/11/2025	11:15	LB134987
	Copper	2.78	2.0	139	-18	22	03/11/2025	11:15	LB134987
	Lead	-10.6			-12	12	03/11/2025	11:15	LB134987
	Nickel	0.028	2.0	1	-38	42	03/11/2025	11:15	LB134987
	Zinc	-177			-40	40	03/11/2025	11:15	LB134987
ICSAB	Cadmium	1020	972	105	826	1120	03/11/2025	11:19	LB134987
	Copper	543	511	106	434	588	03/11/2025	11:19	LB134987
	Lead	71.4	49.0	146	37	61	03/11/2025	11:19	LB134987
	Nickel	1020	954	107	810	1100	03/11/2025	11:19	LB134987
	Zinc	906	952	95	809	1095	03/11/2025	11:19	LB134987
ICSA01	Cadmium	-3.11	1.0	311	-5	7	03/13/2025	12:25	LB135035
	Copper	10.4	2.0	519	-18	22	03/13/2025	12:25	LB135035
	Lead	4.72			-12	12	03/13/2025	12:25	LB135035
	Nickel	1.93	2.0	96	-38	42	03/13/2025	12:25	LB135035
	Zinc	4.65			-40	40	03/13/2025	12:25	LB135035
ICSAB01	Cadmium	1000	972	103	826	1120	03/13/2025	12:30	LB135035
	Copper	504	511	99	434	588	03/13/2025	12:30	LB135035
	Lead	54.2	49.0	111	37	61	03/13/2025	12:30	LB135035
	Nickel	1010	954	106	810	1100	03/13/2025	12:30	LB135035
	Zinc	1050	952	110	809	1095	03/13/2025	12:30	LB135035
ICSA	Cadmium	-2.02	1.0	202	-5	7	03/13/2025	12:34	LB135035
	Copper	5.12	2.0	256	-18	22	03/13/2025	12:34	LB135035
	Lead	1.40			-12	12	03/13/2025	12:34	LB135035
	Nickel	2.27	2.0	114	-38	42	03/13/2025	12:34	LB135035
	Zinc	-9.80			-40	40	03/13/2025	12:34	LB135035
ICSAB	Cadmium	1050	972	108	826	1120	03/13/2025	12:38	LB135035
	Copper	569	511	111	434	588	03/13/2025	12:38	LB135035
	Lead	55.2	49.0	113	37	61	03/13/2025	12:38	LB135035
	Nickel	1050	954	110	810	1100	03/13/2025	12:38	LB135035
	Zinc	1110	952	117	809	1095	03/13/2025	12:38	LB135035

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** Q1491
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1491 **SAS No.:** Q1491
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-4.02	1.0	402	-5	7	03/17/2025	14:51	LB135063
	Copper	17.5	2.0	875	-18	22	03/17/2025	14:51	LB135063
	Lead	7.22			-12	12	03/17/2025	14:51	LB135063
	Nickel	2.20	2.0	110	-38	42	03/17/2025	14:51	LB135063
	Zinc	3.55			-40	40	03/17/2025	14:51	LB135063
ICSAB01	Cadmium	1030	972	106	826	1120	03/17/2025	15:02	LB135063
	Copper	534	511	104	434	588	03/17/2025	15:02	LB135063
	Lead	58.7	49.0	120	37	61	03/17/2025	15:02	LB135063
	Nickel	1030	954	108	810	1100	03/17/2025	15:02	LB135063
	Zinc	827	952	87	809	1095	03/17/2025	15:02	LB135063
ICSA	Cadmium	-5.42	1.0	542	-5	7	03/17/2025	15:06	LB135063
	Copper	16.6	2.0	830	-18	22	03/17/2025	15:06	LB135063
	Lead	2.47			-12	12	03/17/2025	15:06	LB135063
	Nickel	5.92	2.0	296	-38	42	03/17/2025	15:06	LB135063
	Zinc	-9.22			-40	40	03/17/2025	15:06	LB135063
ICSAB	Cadmium	1020	972	105	826	1120	03/17/2025	15:11	LB135063
	Copper	556	511	109	434	588	03/17/2025	15:11	LB135063
	Lead	65.6	49.0	134	37	61	03/17/2025	15:11	LB135063
	Nickel	1020	954	107	810	1100	03/17/2025	15:11	LB135063
	Zinc	1100	952	116	809	1095	03/17/2025	15:11	LB135063