

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1492 SAS No.: Q1492 SDG NO.: Q1492

EPA Sample No.: SSTDICCC040 Date Analyzed: 03/10/2025

Lab File ID: BF141901.D Time Analyzed: 12:58

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	222904	6.881	874166	8.16	493744	9.92
UPPER LIMIT	445808	7.381	1748330	8.663	987488	10.416
LOWER LIMIT	111452	6.381	437083	7.663	246872	9.416
EPA SAMPLE NO.						
01 PB167012BL	215606	6.88	850637	8.16	501095	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.