

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1492 SAS No.: Q1492 SDG NO.: Q1492

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/11/2025

Lab File ID: BF141910.D Time Analyzed: 11:51

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	212930	6.881	840081	8.16	479107	9.92
UPPER LIMIT	425860	7.381	1680160	8.663	958214	10.416
LOWER LIMIT	106465	6.381	420041	7.663	239554	9.416
EPA SAMPLE NO.						
01 PB167012BS	184657	6.88	747989	8.16	446834	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.