

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1492 SAS No.: Q1492 SDG NO.: Q1492

EPA Sample No.: SSTDICCC040 Date Analyzed: 03/10/2025

Lab File ID: BF141901.D Time Analyzed: 12:58

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	839154	11.404	569865	14.039	469286	15.509
UPPER LIMIT	1678310	11.904	1139730	14.539	938572	16.009
LOWER LIMIT	419577	10.904	284933	13.539	234643	15.009
EPA SAMPLE NO.						
01 PB167012BL	952421	11.40	661459	14.03	462716	15.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.