

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1495
Project:	Transfer Station-SPDES	Sample ID:	Q1494-01
Client ID:	PURGE-WATERMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	1.30		0.24		1	1	106		03/06/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1495
Project:	Transfer Station-SPDES	Sample ID:	Q1494-01
Client ID:	PURGE-WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	1.20		0.24		1	1	96		03/06/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1495
Project:	Transfer Station-SPDES	Sample ID:	Q1519-02
Client ID:	WATER TREATMENT DISCHARGEMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	24.6		4.60	J	20.0	1	100		03/07/2025

Matrix Spike Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1495
Project:	Transfer Station-SPDES	Sample ID:	Q1519-02
Client ID:	WATER TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	24.8		4.60	J	20.0	1	101		03/07/2025