

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1495

Project: Transfer Station-SPDES

Run No.: LB134910

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134910BS							
BOD5	mg/L	198	185		93	1	84.6-115.4	03/05/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1495

Project: Transfer Station-SPDES

Run No.: LB134923

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134923BS							
TSS	mg/L	550	533		97	1	90-110	03/06/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1495

Project: Transfer Station-SPDES

Run No.: LB134949

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134949BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	03/07/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1495

Project: Transfer Station-SPDES

Run No.: LB134920

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166997BS							
Ammonia as N	mg/L	1	0.99		99	1	90-110	03/06/2025