

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1040	1000	104	95 - 105	P	03/11/2025	10:50	LB134987
	Cadmium	511	510	100	95 - 105	P	03/11/2025	10:50	LB134987
	Chromium	527	520	101	95 - 105	P	03/11/2025	10:50	LB134987
	Copper	532	510	104	95 - 105	P	03/11/2025	10:50	LB134987
	Iron	10200	10000	102	95 - 105	P	03/11/2025	10:50	LB134987
	Lead	1020	1000	102	95 - 105	P	03/11/2025	10:50	LB134987
	Selenium	1040	1000	104	95 - 105	P	03/11/2025	10:50	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.2	20.0	106	80 - 120	P	03/11/2025	10:54	LB134987
	Cadmium	5.80	6.0	97	80 - 120	P	03/11/2025	10:54	LB134987
	Chromium	10.2	10.0	102	80 - 120	P	03/11/2025	10:54	LB134987
	Copper	22.1	20.0	110	80 - 120	P	03/11/2025	10:54	LB134987
	Iron	93.9	100	94	80 - 120	P	03/11/2025	10:54	LB134987
	Lead	12.0	12.0	100	80 - 120	P	03/11/2025	10:54	LB134987
	Selenium	20.5	20.0	102	80 - 120	P	03/11/2025	10:54	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495

Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5010	5000	100	90 - 110	P	03/11/2025	11:29	LB134987
	Cadmium	2520	2500	101	90 - 110	P	03/11/2025	11:29	LB134987
	Chromium	1010	1000	101	90 - 110	P	03/11/2025	11:29	LB134987
	Copper	1280	1250	102	90 - 110	P	03/11/2025	11:29	LB134987
	Iron	5080	5000	102	90 - 110	P	03/11/2025	11:29	LB134987
	Lead	5030	5000	101	90 - 110	P	03/11/2025	11:29	LB134987
	Selenium	4990	5000	100	90 - 110	P	03/11/2025	11:29	LB134987
CCV02	Arsenic	5150	5000	103	90 - 110	P	03/11/2025	12:19	LB134987
	Cadmium	2570	2500	103	90 - 110	P	03/11/2025	12:19	LB134987
	Chromium	1040	1000	104	90 - 110	P	03/11/2025	12:19	LB134987
	Copper	1300	1250	104	90 - 110	P	03/11/2025	12:19	LB134987
	Iron	4920	5000	98	90 - 110	P	03/11/2025	12:19	LB134987
	Lead	5120	5000	102	90 - 110	P	03/11/2025	12:19	LB134987
	Selenium	5170	5000	103	90 - 110	P	03/11/2025	12:19	LB134987
CCV03	Arsenic	4860	5000	97	90 - 110	P	03/11/2025	13:10	LB134987
	Cadmium	2430	2500	97	90 - 110	P	03/11/2025	13:10	LB134987
	Chromium	964	1000	96	90 - 110	P	03/11/2025	13:10	LB134987
	Copper	1250	1250	100	90 - 110	P	03/11/2025	13:10	LB134987
	Iron	4750	5000	95	90 - 110	P	03/11/2025	13:10	LB134987
	Lead	4830	5000	97	90 - 110	P	03/11/2025	13:10	LB134987
	Selenium	4840	5000	97	90 - 110	P	03/11/2025	13:10	LB134987
CCV04	Arsenic	4940	5000	99	90 - 110	P	03/11/2025	14:08	LB134987
	Cadmium	2470	2500	99	90 - 110	P	03/11/2025	14:08	LB134987
	Chromium	965	1000	96	90 - 110	P	03/11/2025	14:08	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	14:08	LB134987
	Iron	4970	5000	100	90 - 110	P	03/11/2025	14:08	LB134987
	Lead	4880	5000	98	90 - 110	P	03/11/2025	14:08	LB134987
	Selenium	4920	5000	98	90 - 110	P	03/11/2025	14:08	LB134987
CCV05	Arsenic	4810	5000	96	90 - 110	P	03/11/2025	15:00	LB134987
	Cadmium	2400	2500	96	90 - 110	P	03/11/2025	15:00	LB134987
	Chromium	934	1000	93	90 - 110	P	03/11/2025	15:00	LB134987
	Copper	1240	1250	99	90 - 110	P	03/11/2025	15:00	LB134987
	Iron	4750	5000	95	90 - 110	P	03/11/2025	15:00	LB134987
	Lead	4760	5000	95	90 - 110	P	03/11/2025	15:00	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495

Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Selenium	4790	5000	96	90 - 110	P	03/11/2025	15:00	LB134987
CCV06	Arsenic	4820	5000	96	90 - 110	P	03/11/2025	15:50	LB134987
	Cadmium	2430	2500	97	90 - 110	P	03/11/2025	15:50	LB134987
	Chromium	956	1000	96	90 - 110	P	03/11/2025	15:50	LB134987
	Copper	1240	1250	99	90 - 110	P	03/11/2025	15:50	LB134987
	Iron	4680	5000	94	90 - 110	P	03/11/2025	15:50	LB134987
	Lead	4800	5000	96	90 - 110	P	03/11/2025	15:50	LB134987
	Selenium	4770	5000	96	90 - 110	P	03/11/2025	15:50	LB134987
CCV07	Arsenic	4920	5000	98	90 - 110	P	03/11/2025	16:42	LB134987
	Cadmium	2460	2500	98	90 - 110	P	03/11/2025	16:42	LB134987
	Chromium	956	1000	96	90 - 110	P	03/11/2025	16:42	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	16:42	LB134987
	Iron	4860	5000	97	90 - 110	P	03/11/2025	16:42	LB134987
	Lead	4850	5000	97	90 - 110	P	03/11/2025	16:42	LB134987
	Selenium	4910	5000	98	90 - 110	P	03/11/2025	16:42	LB134987
CCV08	Arsenic	4940	5000	99	90 - 110	P	03/11/2025	17:34	LB134987
	Cadmium	2480	2500	99	90 - 110	P	03/11/2025	17:34	LB134987
	Chromium	955	1000	96	90 - 110	P	03/11/2025	17:34	LB134987
	Copper	1270	1250	102	90 - 110	P	03/11/2025	17:34	LB134987
	Iron	4880	5000	98	90 - 110	P	03/11/2025	17:34	LB134987
	Lead	4880	5000	98	90 - 110	P	03/11/2025	17:34	LB134987
	Selenium	4930	5000	99	90 - 110	P	03/11/2025	17:34	LB134987
CCV09	Arsenic	5030	5000	101	90 - 110	P	03/11/2025	18:26	LB134987
	Cadmium	2520	2500	101	90 - 110	P	03/11/2025	18:26	LB134987
	Chromium	971	1000	97	90 - 110	P	03/11/2025	18:26	LB134987
	Copper	1290	1250	103	90 - 110	P	03/11/2025	18:26	LB134987
	Iron	4690	5000	94	90 - 110	P	03/11/2025	18:26	LB134987
	Lead	4960	5000	99	90 - 110	P	03/11/2025	18:26	LB134987
	Selenium	5030	5000	101	90 - 110	P	03/11/2025	18:26	LB134987
CCV10	Arsenic	4890	5000	98	90 - 110	P	03/11/2025	19:14	LB134987
	Cadmium	2480	2500	99	90 - 110	P	03/11/2025	19:14	LB134987
	Chromium	968	1000	97	90 - 110	P	03/11/2025	19:14	LB134987
	Copper	1260	1250	101	90 - 110	P	03/11/2025	19:14	LB134987
	Iron	5280	5000	106	90 - 110	P	03/11/2025	19:14	LB134987

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc SDG No.: Q1495
Contract: TULL01 Lab Code: CHEM Case No.: Q1495 SAS No.: Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Lead	4880	5000	98	90 - 110	P	03/11/2025	19:14	LB134987
	Selenium	4860	5000	97	90 - 110	P	03/11/2025	19:14	LB134987
CCV11	Arsenic	4830	5000	97	90 - 110	P	03/11/2025	19:26	LB134987
	Cadmium	2390	2500	95	90 - 110	P	03/11/2025	19:26	LB134987
	Chromium	935	1000	94	90 - 110	P	03/11/2025	19:26	LB134987
	Copper	1240	1250	100	90 - 110	P	03/11/2025	19:26	LB134987
	Iron	4770	5000	95	90 - 110	P	03/11/2025	19:26	LB134987
	Lead	4740	5000	95	90 - 110	P	03/11/2025	19:26	LB134987
	Selenium	4840	5000	97	90 - 110	P	03/11/2025	19:26	LB134987
CCV12	Arsenic	4850	5000	97	90 - 110	P	03/11/2025	19:39	LB134987
	Cadmium	2380	2500	95	90 - 110	P	03/11/2025	19:39	LB134987
	Chromium	948	1000	95	90 - 110	P	03/11/2025	19:39	LB134987
	Copper	1230	1250	99	90 - 110	P	03/11/2025	19:39	LB134987
	Iron	4570	5000	91	90 - 110	P	03/11/2025	19:39	LB134987
	Lead	4680	5000	94	90 - 110	P	03/11/2025	19:39	LB134987
	Selenium	5040	5000	101	90 - 110	P	03/11/2025	19:39	LB134987



- 2a -

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q1495</u>	
Contract:	<u>TULL01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1495</u> SAS No.: <u>Q1495</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
ICV01	Arsenic	1020	1000	102	95 - 105	P	03/13/2025	12:07	LB135035
	Cadmium	503	510	99	95 - 105	P	03/13/2025	12:07	LB135035
	Chromium	525	520	101	95 - 105	P	03/13/2025	12:07	LB135035
	Copper	529	510	104	95 - 105	P	03/13/2025	12:07	LB135035
	Iron	9860	10000	99	95 - 105	P	03/13/2025	12:07	LB135035
	Lead	1000	1000	100	95 - 105	P	03/13/2025	12:07	LB135035
	Selenium	1040	1000	104	95 - 105	P	03/13/2025	12:07	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	17.3	20.0	86	80 - 120	P	03/13/2025	12:13	LB135035
	Cadmium	5.95	6.0	99	80 - 120	P	03/13/2025	12:13	LB135035
	Chromium	10.1	10.0	101	80 - 120	P	03/13/2025	12:13	LB135035
	Copper	22.7	20.0	114	80 - 120	P	03/13/2025	12:13	LB135035
	Iron	101	100	101	80 - 120	P	03/13/2025	12:13	LB135035
	Lead	12.6	12.0	105	80 - 120	P	03/13/2025	12:13	LB135035
	Selenium	19.2	20.0	96	80 - 120	P	03/13/2025	12:13	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5160	5000	103	90 - 110	P	03/13/2025	12:53	LB135035
	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	12:53	LB135035
	Chromium	1050	1000	105	90 - 110	P	03/13/2025	12:53	LB135035
	Copper	1280	1250	103	90 - 110	P	03/13/2025	12:53	LB135035
	Iron	4890	5000	98	90 - 110	P	03/13/2025	12:53	LB135035
	Lead	5190	5000	104	90 - 110	P	03/13/2025	12:53	LB135035
	Selenium	5070	5000	101	90 - 110	P	03/13/2025	12:53	LB135035
CCV02	Arsenic	5290	5000	106	90 - 110	P	03/13/2025	13:49	LB135035
	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	13:49	LB135035
	Chromium	1100	1000	110	90 - 110	P	03/13/2025	13:49	LB135035
	Copper	1310	1250	105	90 - 110	P	03/13/2025	13:49	LB135035
	Iron	5130	5000	103	90 - 110	P	03/13/2025	13:49	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	13:49	LB135035
	Selenium	5180	5000	104	90 - 110	P	03/13/2025	13:49	LB135035
CCV03	Arsenic	5130	5000	102	90 - 110	P	03/13/2025	14:40	LB135035
	Cadmium	2540	2500	102	90 - 110	P	03/13/2025	14:40	LB135035
	Chromium	1050	1000	106	90 - 110	P	03/13/2025	14:40	LB135035
	Copper	1270	1250	102	90 - 110	P	03/13/2025	14:40	LB135035
	Iron	4990	5000	100	90 - 110	P	03/13/2025	14:40	LB135035
	Lead	5150	5000	103	90 - 110	P	03/13/2025	14:40	LB135035
	Selenium	5060	5000	101	90 - 110	P	03/13/2025	14:40	LB135035
CCV04	Arsenic	4950	5000	99	90 - 110	P	03/13/2025	15:38	LB135035
	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	15:38	LB135035
	Chromium	1030	1000	103	90 - 110	P	03/13/2025	15:38	LB135035
	Copper	1240	1250	99	90 - 110	P	03/13/2025	15:38	LB135035
	Iron	4780	5000	96	90 - 110	P	03/13/2025	15:38	LB135035
	Lead	5050	5000	101	90 - 110	P	03/13/2025	15:38	LB135035
	Selenium	4830	5000	97	90 - 110	P	03/13/2025	15:38	LB135035
CCV05	Arsenic	5250	5000	105	90 - 110	P	03/13/2025	16:28	LB135035
	Cadmium	2630	2500	105	90 - 110	P	03/13/2025	16:28	LB135035
	Chromium	1070	1000	107	90 - 110	P	03/13/2025	16:28	LB135035
	Copper	1300	1250	104	90 - 110	P	03/13/2025	16:28	LB135035
	Iron	5120	5000	102	90 - 110	P	03/13/2025	16:28	LB135035
	Lead	5300	5000	106	90 - 110	P	03/13/2025	16:28	LB135035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Selenium	5150	5000	103	90 - 110	P	03/13/2025	16:28	LB135035
CCV06	Arsenic	5070	5000	101	90 - 110	P	03/13/2025	17:21	LB135035
	Cadmium	2570	2500	103	90 - 110	P	03/13/2025	17:21	LB135035
	Chromium	1060	1000	106	90 - 110	P	03/13/2025	17:21	LB135035
	Copper	1270	1250	102	90 - 110	P	03/13/2025	17:21	LB135035
	Iron	5040	5000	101	90 - 110	P	03/13/2025	17:21	LB135035
	Lead	5180	5000	104	90 - 110	P	03/13/2025	17:21	LB135035
	Selenium	4970	5000	100	90 - 110	P	03/13/2025	17:21	LB135035
CCV07	Arsenic	5250	5000	105	90 - 110	P	03/13/2025	18:12	LB135035
	Cadmium	2640	2500	106	90 - 110	P	03/13/2025	18:12	LB135035
	Chromium	1090	1000	109	90 - 110	P	03/13/2025	18:12	LB135035
	Copper	1310	1250	105	90 - 110	P	03/13/2025	18:12	LB135035
	Iron	5150	5000	103	90 - 110	P	03/13/2025	18:12	LB135035
	Lead	5330	5000	107	90 - 110	P	03/13/2025	18:12	LB135035
	Selenium	5180	5000	104	90 - 110	P	03/13/2025	18:12	LB135035
CCV08	Arsenic	4950	5000	99	90 - 110	P	03/13/2025	18:53	LB135035
	Cadmium	2510	2500	100	90 - 110	P	03/13/2025	18:53	LB135035
	Chromium	1040	1000	104	90 - 110	P	03/13/2025	18:53	LB135035
	Copper	1240	1250	99	90 - 110	P	03/13/2025	18:53	LB135035
	Iron	4880	5000	98	90 - 110	P	03/13/2025	18:53	LB135035
	Lead	5060	5000	101	90 - 110	P	03/13/2025	18:53	LB135035
	Selenium	4840	5000	97	90 - 110	P	03/13/2025	18:53	LB135035
CCV09	Arsenic	5280	5000	106	90 - 110	P	03/13/2025	19:12	LB135035
	Cadmium	2550	2500	102	90 - 110	P	03/13/2025	19:12	LB135035
	Chromium	1070	1000	107	90 - 110	P	03/13/2025	19:12	LB135035
	Copper	1290	1250	103	90 - 110	P	03/13/2025	19:12	LB135035
	Iron	5230	5000	105	90 - 110	P	03/13/2025	19:12	LB135035
	Lead	5120	5000	102	90 - 110	P	03/13/2025	19:12	LB135035
	Selenium	5440	5000	109	90 - 110	P	03/13/2025	19:12	LB135035



- 2a -

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q1495</u>	
Contract:	<u>TULL01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1495</u> SAS No.: <u>Q1495</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
ICV01	Arsenic	1020	1000	102	95 - 105	P	03/17/2025	14:31	LB135063
	Cadmium	504	510	99	95 - 105	P	03/17/2025	14:31	LB135063
	Chromium	529	520	102	95 - 105	P	03/17/2025	14:31	LB135063
	Copper	528	510	104	95 - 105	P	03/17/2025	14:31	LB135063
	Iron	10100	10000	101	95 - 105	P	03/17/2025	14:31	LB135063
	Lead	996	1000	100	95 - 105	P	03/17/2025	14:31	LB135063
	Selenium	1050	1000	105	95 - 105	P	03/17/2025	14:31	LB135063

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.4	20.0	97	80 - 120	P	03/17/2025	14:38	LB135063
	Cadmium	5.89	6.0	98	80 - 120	P	03/17/2025	14:38	LB135063
	Chromium	10.3	10.0	103	80 - 120	P	03/17/2025	14:38	LB135063
	Copper	22.1	20.0	110	80 - 120	P	03/17/2025	14:38	LB135063
	Iron	99.8	100	100	80 - 120	P	03/17/2025	14:38	LB135063
	Lead	12.3	12.0	103	80 - 120	P	03/17/2025	14:38	LB135063
	Selenium	18.2	20.0	91	80 - 120	P	03/17/2025	14:38	LB135063

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495

Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5160	5000	103	90 - 110	P	03/17/2025	15:15	LB135063
	Cadmium	2510	2500	100	90 - 110	P	03/17/2025	15:15	LB135063
	Chromium	1030	1000	103	90 - 110	P	03/17/2025	15:15	LB135063
	Copper	1270	1250	102	90 - 110	P	03/17/2025	15:15	LB135063
	Iron	5180	5000	104	90 - 110	P	03/17/2025	15:15	LB135063
	Lead	5080	5000	102	90 - 110	P	03/17/2025	15:15	LB135063
	Selenium	5080	5000	102	90 - 110	P	03/17/2025	15:15	LB135063
CCV02	Arsenic	5220	5000	104	90 - 110	P	03/17/2025	16:13	LB135063
	Cadmium	2500	2500	100	90 - 110	P	03/17/2025	16:13	LB135063
	Chromium	1040	1000	104	90 - 110	P	03/17/2025	16:13	LB135063
	Copper	1280	1250	102	90 - 110	P	03/17/2025	16:13	LB135063
	Iron	5090	5000	102	90 - 110	P	03/17/2025	16:13	LB135063
	Lead	5070	5000	101	90 - 110	P	03/17/2025	16:13	LB135063
	Selenium	5140	5000	103	90 - 110	P	03/17/2025	16:13	LB135063
CCV03	Arsenic	5320	5000	106	90 - 110	P	03/17/2025	17:05	LB135063
	Cadmium	2570	2500	103	90 - 110	P	03/17/2025	17:05	LB135063
	Chromium	1090	1000	109	90 - 110	P	03/17/2025	17:05	LB135063
	Copper	1310	1250	104	90 - 110	P	03/17/2025	17:05	LB135063
	Iron	5130	5000	103	90 - 110	P	03/17/2025	17:05	LB135063
	Lead	5200	5000	104	90 - 110	P	03/17/2025	17:05	LB135063
	Selenium	5220	5000	104	90 - 110	P	03/17/2025	17:05	LB135063
CCV04	Arsenic	5020	5000	100	90 - 110	P	03/17/2025	17:51	LB135063
	Cadmium	2390	2500	96	90 - 110	P	03/17/2025	17:51	LB135063
	Chromium	1010	1000	101	90 - 110	P	03/17/2025	17:51	LB135063
	Copper	1230	1250	98	90 - 110	P	03/17/2025	17:51	LB135063
	Iron	4880	5000	98	90 - 110	P	03/17/2025	17:51	LB135063
	Lead	4870	5000	97	90 - 110	P	03/17/2025	17:51	LB135063
	Selenium	4900	5000	98	90 - 110	P	03/17/2025	17:51	LB135063
CCV05	Arsenic	5050	5000	101	90 - 110	P	03/17/2025	18:18	LB135063
	Cadmium	2390	2500	96	90 - 110	P	03/17/2025	18:18	LB135063
	Chromium	1020	1000	102	90 - 110	P	03/17/2025	18:18	LB135063
	Copper	1240	1250	99	90 - 110	P	03/17/2025	18:18	LB135063
	Iron	4960	5000	99	90 - 110	P	03/17/2025	18:18	LB135063
	Lead	4870	5000	97	90 - 110	P	03/17/2025	18:18	LB135063

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Selenium	4960	5000	99	90 - 110	P	03/17/2025	18:18	LB135063
