



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc SDG No.: Q1495
Contract: TULL01 Lab Code: CHEM Case No.: Q1495 SAS No.: Q1495
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.0	20.0	95	40 - 160	P	03/11/2025	11:02	LB134987
	Cadmium	5.81	6.0	97	40 - 160	P	03/11/2025	11:02	LB134987
	Chromium	9.83	10.0	98	40 - 160	P	03/11/2025	11:02	LB134987
	Copper	22.1	20.0	110	40 - 160	P	03/11/2025	11:02	LB134987
	Iron	89.9	100	90	40 - 160	P	03/11/2025	11:02	LB134987
	Lead	11.9	12.0	99	40 - 160	P	03/11/2025	11:02	LB134987
	Selenium	21.9	20.0	110	40 - 160	P	03/11/2025	11:02	LB134987
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	03/13/2025	12:21	LB135035
	Cadmium	5.90	6.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Chromium	10.4	10.0	104	40 - 160	P	03/13/2025	12:21	LB135035
	Copper	23.1	20.0	116	40 - 160	P	03/13/2025	12:21	LB135035
	Iron	110	100	110	40 - 160	P	03/13/2025	12:21	LB135035
	Lead	11.7	12.0	98	40 - 160	P	03/13/2025	12:21	LB135035
	Selenium	21.3	20.0	107	40 - 160	P	03/13/2025	12:21	LB135035
CRI01	Arsenic	19.6	20.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Cadmium	5.88	6.0	98	40 - 160	P	03/17/2025	14:47	LB135063
	Chromium	10.3	10.0	103	40 - 160	P	03/17/2025	14:47	LB135063
	Copper	22.2	20.0	111	40 - 160	P	03/17/2025	14:47	LB135063
	Iron	98.2	100	98	40 - 160	P	03/17/2025	14:47	LB135063
	Lead	12.6	12.0	105	40 - 160	P	03/17/2025	14:47	LB135063
	Selenium	18.9	20.0	94	40 - 160	P	03/17/2025	14:47	LB135063