



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q1495

Contract: TULL01

Lab Code: CHEM

Case No.: Q1495

SAS No.: Q1495

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	10:58	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	10:58	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	10:58	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	10:58	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	10:58	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	10:58	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	10:58	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1495							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1495	SAS No.: Q1495						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	11:34	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	11:34	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	11:34	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	11:34	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	11:34	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	11:34	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	11:34	LB134987
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	12:23	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	12:23	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	12:23	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	12:23	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	12:23	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	12:23	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	12:23	LB134987
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	13:14	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	13:14	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	13:14	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	13:14	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	13:14	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	13:14	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	13:14	LB134987
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	14:13	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	14:13	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	14:13	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	14:13	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	14:13	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	14:13	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	14:13	LB134987
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	15:04	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	15:04	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	15:04	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	15:04	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	15:04	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	15:04	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	15:04	LB134987
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	15:54	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	15:54	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	15:54	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	15:54	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1495							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1495	SAS No.: Q1495						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Iron	100	+/-100	U	100	P	03/11/2025	15:54	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	15:54	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	15:54	LB134987
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	16:46	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	16:46	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	16:46	LB134987
CCB08	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	16:46	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	16:46	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	16:46	LB134987
CCB09	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	16:46	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	17:38	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	17:38	LB134987
CCB10	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	17:38	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	17:38	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	17:38	LB134987
CCB11	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	17:38	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	17:38	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	18:30	LB134987
CCB12	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	18:30	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	18:30	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	18:30	LB134987
CCB13	Iron	100	+/-100	U	100	P	03/11/2025	18:30	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	18:30	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	18:30	LB134987
CCB14	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:18	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:18	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:18	LB134987
CCB15	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	19:18	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	19:18	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:18	LB134987
CCB16	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:18	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:31	LB134987
	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:31	LB134987
CCB17	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:31	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	19:31	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	19:31	LB134987
CCB18	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:31	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:31	LB134987
	Arsenic	20.0	+/-20.0	U	20.0	P	03/11/2025	19:49	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Cadmium	6.00	+/-6.00	U	6.00	P	03/11/2025	19:49	LB134987
	Chromium	10.0	+/-10.0	U	10.0	P	03/11/2025	19:49	LB134987
	Copper	20.0	+/-20.0	U	20.0	P	03/11/2025	19:49	LB134987
	Iron	100	+/-100	U	100	P	03/11/2025	19:49	LB134987
	Lead	12.0	+/-12.0	U	12.0	P	03/11/2025	19:49	LB134987
	Selenium	20.0	+/-20.0	U	20.0	P	03/11/2025	19:49	LB134987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>Q1495</u>							
Contract:	<u>TULL01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1495</u>	SAS No.:	<u>Q1495</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:17	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	12:17	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	12:17	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	12:17	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	12:17	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1495							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1495	SAS No.: Q1495						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	12:58	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	12:58	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	12:58	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	12:58	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	12:58	LB135035
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	13:54	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	13:54	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	13:54	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	13:54	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	13:54	LB135035
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	14:48	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	14:48	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	14:48	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	14:48	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	14:48	LB135035
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	15:42	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	15:42	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	15:42	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	15:42	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	15:42	LB135035
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	16:32	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	16:32	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	16:32	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	16:32	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	16:32	LB135035
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	17:25	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	17:25	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>Q1495</u>							
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1495</u>	SAS No.: <u>Q1495</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Iron	100	+/-100	U	100	P	03/13/2025	17:25	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	17:25	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	17:25	LB135035
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	18:35	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	18:35	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Iron	45.6	+/-100	J	100	P	03/13/2025	18:35	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	18:35	LB135035
CCB08	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	18:35	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:03	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	19:03	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	19:03	LB135035
CCB09	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	19:03	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	19:03	LB135035
	Arsenic	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Cadmium	6.00	+/-6.00	U	6.00	P	03/13/2025	19:21	LB135035
	Chromium	10.0	+/-10.0	U	10.0	P	03/13/2025	19:21	LB135035
	Copper	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035
	Iron	100	+/-100	U	100	P	03/13/2025	19:21	LB135035
	Lead	12.0	+/-12.0	U	12.0	P	03/13/2025	19:21	LB135035
	Selenium	20.0	+/-20.0	U	20.0	P	03/13/2025	19:21	LB135035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>Q1495</u>							
Contract: <u>TULL01</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1495</u>		SAS No.: <u>Q1495</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	14:43	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	14:43	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	14:43	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	14:43	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	14:43	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	14:43	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	14:43	LB135063

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc		SDG No.: Q1495							
Contract: TULL01	Lab Code: CHEM	Case No.: Q1495	SAS No.: Q1495						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	15:27	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	15:27	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	15:27	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	15:27	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	15:27	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	15:27	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	15:27	LB135063
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	16:17	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	16:17	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	16:17	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	16:17	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	16:17	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	16:17	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	16:17	LB135063
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	17:09	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	17:09	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	17:09	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	17:09	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	17:09	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	17:09	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	17:09	LB135063
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	17:55	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	17:55	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	17:55	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	17:55	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	17:55	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	17:55	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	17:55	LB135063
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/17/2025	18:22	LB135063
	Cadmium	6.00	+/-6.00	U	6.00	P	03/17/2025	18:22	LB135063
	Chromium	10.0	+/-10.0	U	10.0	P	03/17/2025	18:22	LB135063
	Copper	20.0	+/-20.0	U	20.0	P	03/17/2025	18:22	LB135063
	Iron	100	+/-100	U	100	P	03/17/2025	18:22	LB135063
	Lead	12.0	+/-12.0	U	12.0	P	03/17/2025	18:22	LB135063
	Selenium	20.0	+/-20.0	U	20.0	P	03/17/2025	18:22	LB135063



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Tully Environmental, Inc</u>			SDG No.:	<u>Q1495</u>		
Contract:	<u>TULL01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1495</u>	SAS No.:	<u>Q1495</u>

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------
