

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	6.46			-20	20	03/11/2025	11:06	LB134987
	Cadmium	-2.31	1.0	231	-5	7	03/11/2025	11:06	LB134987
	Chromium	58.2	52.0	112	42	62	03/11/2025	11:06	LB134987
	Copper	6.76	2.0	338	-18	22	03/11/2025	11:06	LB134987
	Iron	102000	101000	101	85600	116500	03/11/2025	11:06	LB134987
	Lead	3.81			-12	12	03/11/2025	11:06	LB134987
	Selenium	-16.3			-20	20	03/11/2025	11:06	LB134987
ICSAB01	Arsenic	114	104	110	88.4	120	03/11/2025	11:11	LB134987
	Cadmium	1020	972	105	826	1120	03/11/2025	11:11	LB134987
	Chromium	567	542	105	460	624	03/11/2025	11:11	LB134987
	Copper	507	511	99	434	588	03/11/2025	11:11	LB134987
	Iron	97400	99300	98	84400	114500	03/11/2025	11:11	LB134987
	Lead	55.7	49.0	114	37	61	03/11/2025	11:11	LB134987
	Selenium	32.4	46.0	70	26	66	03/11/2025	11:11	LB134987
ICSA	Arsenic	-14.8			-20	20	03/11/2025	11:15	LB134987
	Cadmium	-2.80	1.0	280	-5	7	03/11/2025	11:15	LB134987
	Chromium	54.3	52.0	104	42	62	03/11/2025	11:15	LB134987
	Copper	2.78	2.0	139	-18	22	03/11/2025	11:15	LB134987
	Iron	101000	101000	100	85600	116500	03/11/2025	11:15	LB134987
	Lead	-10.6			-12	12	03/11/2025	11:15	LB134987
	Selenium	11.4			-20	20	03/11/2025	11:15	LB134987
ICSAB	Arsenic	77.1	104	74	88.4	120	03/11/2025	11:19	LB134987
	Cadmium	1020	972	105	826	1120	03/11/2025	11:19	LB134987
	Chromium	596	542	110	460	624	03/11/2025	11:19	LB134987
	Copper	543	511	106	434	588	03/11/2025	11:19	LB134987
	Iron	96400	99300	97	84400	114500	03/11/2025	11:19	LB134987
	Lead	71.4	49.0	146	37	61	03/11/2025	11:19	LB134987
	Selenium	32.7	46.0	71	26	66	03/11/2025	11:19	LB134987
ICSA01	Arsenic	4.32			-20	20	03/13/2025	12:25	LB135035
	Cadmium	-3.11	1.0	311	-5	7	03/13/2025	12:25	LB135035
	Chromium	60.0	52.0	115	42	62	03/13/2025	12:25	LB135035
	Copper	10.4	2.0	519	-18	22	03/13/2025	12:25	LB135035
	Iron	103000	101000	102	85600	116500	03/13/2025	12:25	LB135035
	Lead	4.72			-12	12	03/13/2025	12:25	LB135035
	Selenium	-15.2			-20	20	03/13/2025	12:25	LB135035
ICSAB01	Arsenic	113	104	109	88.4	120	03/13/2025	12:30	LB135035
	Cadmium	1000	972	103	826	1120	03/13/2025	12:30	LB135035
	Chromium	564	542	104	460	624	03/13/2025	12:30	LB135035
	Copper	504	511	99	434	588	03/13/2025	12:30	LB135035
	Iron	100000	99300	101	84400	114500	03/13/2025	12:30	LB135035
	Lead	54.2	49.0	111	37	61	03/13/2025	12:30	LB135035

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tully Environmental, Inc</u>	SDG No.: <u>Q1495</u>
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1495</u>
	SAS No.: <u>Q1495</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	43.0	46.0	94	26	66	03/13/2025	12:30	LB135035
ICSA	Arsenic	-26.0			-20	20	03/13/2025	12:34	LB135035
	Cadmium	-2.02	1.0	202	-5	7	03/13/2025	12:34	LB135035
	Chromium	53.3	52.0	102	42	62	03/13/2025	12:34	LB135035
	Copper	5.12	2.0	256	-18	22	03/13/2025	12:34	LB135035
	Iron	92100	101000	91	85600	116500	03/13/2025	12:34	LB135035
	Lead	1.40			-12	12	03/13/2025	12:34	LB135035
	Selenium	42.2			-20	20	03/13/2025	12:34	LB135035
ICSAB	Arsenic	110	104	106	88.4	120	03/13/2025	12:38	LB135035
	Cadmium	1050	972	108	826	1120	03/13/2025	12:38	LB135035
	Chromium	631	542	116	460	624	03/13/2025	12:38	LB135035
	Copper	569	511	111	434	588	03/13/2025	12:38	LB135035
	Iron	101000	99300	102	84400	114500	03/13/2025	12:38	LB135035
	Lead	55.2	49.0	113	37	61	03/13/2025	12:38	LB135035
	Selenium	109	46.0	237	26	66	03/13/2025	12:38	LB135035
ICSA01	Arsenic	4.68			-20	20	03/17/2025	14:51	LB135063
	Cadmium	-4.02	1.0	402	-5	7	03/17/2025	14:51	LB135063
	Chromium	60.2	52.0	116	42	62	03/17/2025	14:51	LB135063
	Copper	17.5	2.0	875	-18	22	03/17/2025	14:51	LB135063
	Iron	101000	101000	100	85600	116500	03/17/2025	14:51	LB135063
	Lead	7.22			-12	12	03/17/2025	14:51	LB135063
	Selenium	-0.91			-20	20	03/17/2025	14:51	LB135063
ICSAB01	Arsenic	95.9	104	92	88.4	120	03/17/2025	15:02	LB135063
	Cadmium	1030	972	106	826	1120	03/17/2025	15:02	LB135063
	Chromium	588	542	108	460	624	03/17/2025	15:02	LB135063
	Copper	534	511	104	434	588	03/17/2025	15:02	LB135063
	Iron	100000	99300	101	84400	114500	03/17/2025	15:02	LB135063
	Lead	58.7	49.0	120	37	61	03/17/2025	15:02	LB135063
	Selenium	57.1	46.0	124	26	66	03/17/2025	15:02	LB135063
ICSA	Arsenic	-2.46			-20	20	03/17/2025	15:06	LB135063
	Cadmium	-5.42	1.0	542	-5	7	03/17/2025	15:06	LB135063
	Chromium	56.1	52.0	108	42	62	03/17/2025	15:06	LB135063
	Copper	16.6	2.0	830	-18	22	03/17/2025	15:06	LB135063
	Iron	101000	101000	100	85600	116500	03/17/2025	15:06	LB135063
	Lead	2.47			-12	12	03/17/2025	15:06	LB135063
	Selenium	-47.6			-20	20	03/17/2025	15:06	LB135063
ICSAB	Arsenic	89.4	104	86	88.4	120	03/17/2025	15:11	LB135063
	Cadmium	1020	972	105	826	1120	03/17/2025	15:11	LB135063
	Chromium	601	542	111	460	624	03/17/2025	15:11	LB135063
	Copper	556	511	109	434	588	03/17/2025	15:11	LB135063

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB	Iron	104000	99300	105	84400	114500	03/17/2025	15:11	LB135063
	Lead	65.6	49.0	134	37	61	03/17/2025	15:11	LB135063
	Selenium	16.0	46.0	35	26	66	03/17/2025	15:11	LB135063