

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1495
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1495 **SAS No.:** Q1495

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167007BS							
Arsenic	ug/L	400	398		100	85 - 115	P
Cadmium	ug/L	100	96.5		96	85 - 115	P
Chromium	ug/L	200	206		103	85 - 115	P
Copper	ug/L	150	156		104	85 - 115	P
Iron	ug/L	1500	1510		101	85 - 115	P
Lead	ug/L	500	482		96	85 - 115	P
Selenium	ug/L	1000	995		100	85 - 115	P