

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1495 SAS No.: Q1495 SDG No.: Q1495

Instrument ID: MSVOA_X Calibration Date/Time: 03/14/2025 10:17

Lab File ID: VX045285.D Init. Calib. Date(s): 02/21/2025 02/21/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:58 11:29

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.731	1.704	0.5	-1.56	
Toluene	1.987	1.976	0.4	-0.55	
Ethyl Benzene	2.176	2.166	0.1	-0.46	
m/p-Xylenes	0.811	0.799	0.3	-1.48	
o-Xylene	0.795	0.783	0.3	-1.51	
1,2-Dichloroethane-d4	2.914	3.149	0.01	8.06	
Toluene-d8	1.660	1.716	0.01	3.37	
4-Bromofluorobenzene	0.561	0.568	0.2	1.25	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.