

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: Q1495 SAS No.: Q1495 SDG NO.: Q1495
Lab File ID: VX045285.D Date Analyzed: 03/14/2025
Instrument ID: MSVOA_X Time Analyzed: 10:17
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	16402	4.89	86549	6.75	77791	10.05
UPPER LIMIT	32804	5.385	173098	7.251	155582	10.549
LOWER LIMIT	8201	4.385	43274.5	6.251	38895.5	9.549
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (MAR)	16787	4.90	89621	6.76	84103	10.06
002-35TH-AVE (MAR)	16144	4.90	84818	6.76	76342	10.06
VX0314WBL01	14558	4.89	75988	6.76	67387	10.06
VX0314WBS01	16027	4.89	88526	6.76	80688	10.05
VX0314WBSD01	15546	4.89	80769	6.75	76870	10.05

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.