

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1498 SAS No.: Q1498 SDG NO.: Q1498

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109648.D Time Analyzed: 10:05

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 4.791 | 4.691                | 4.891 | 495.490            | 500.000           | -0.9  |
| Aroclor-1016-2       | 4.811 | 4.711                | 4.911 | 499.000            | 500.000           | -0.2  |
| Aroclor-1016-3       | 4.867 | 4.767                | 4.967 | 498.220            | 500.000           | -0.4  |
| Aroclor-1016-4       | 4.988 | 4.888                | 5.088 | 499.640            | 500.000           | -0.1  |
| Aroclor-1016-5       | 5.244 | 5.145                | 5.345 | 479.950            | 500.000           | -4.0  |
| Aroclor-1260-1       | 6.288 | 6.188                | 6.388 | 497.990            | 500.000           | -0.4  |
| Aroclor-1260-2       | 6.476 | 6.377                | 6.577 | 491.040            | 500.000           | -1.8  |
| Aroclor-1260-3       | 6.846 | 6.746                | 6.946 | 477.970            | 500.000           | -4.4  |
| Aroclor-1260-4       | 7.106 | 7.005                | 7.205 | 481.150            | 500.000           | -3.8  |
| Aroclor-1260-5       | 7.349 | 7.247                | 7.447 | 485.050            | 500.000           | -3.0  |
| Decachlorobiphenyl   | 8.757 | 8.657                | 8.857 | 43.950             | 50.000            | -12.1 |
| Tetrachloro-m-xylene | 3.697 | 3.598                | 3.798 | 48.860             | 50.000            | -2.3  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1498 SAS No.: Q1498 SDG NO.: Q1498

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/20/2025 02/20/2025

Client Sample No.: CCAL01 Date Analyzed: 03/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO109648.D Time Analyzed: 10:05

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.778 | 4.678                | 4.878 | 524.970            | 500.000           | 5.0  |
| Aroclor-1016-2       | 4.797 | 4.697                | 4.897 | 531.650            | 500.000           | 6.3  |
| Aroclor-1016-3       | 4.972 | 4.872                | 5.072 | 532.380            | 500.000           | 6.5  |
| Aroclor-1016-4       | 5.014 | 4.914                | 5.114 | 451.780            | 500.000           | -9.6 |
| Aroclor-1016-5       | 5.227 | 5.127                | 5.327 | 504.620            | 500.000           | 0.9  |
| Aroclor-1260-1       | 6.261 | 6.161                | 6.361 | 517.600            | 500.000           | 3.5  |
| Aroclor-1260-2       | 6.449 | 6.348                | 6.548 | 516.800            | 500.000           | 3.4  |
| Aroclor-1260-3       | 6.602 | 6.502                | 6.702 | 505.870            | 500.000           | 1.2  |
| Aroclor-1260-4       | 7.074 | 6.974                | 7.174 | 499.390            | 500.000           | -0.1 |
| Aroclor-1260-5       | 7.315 | 7.214                | 7.414 | 511.830            | 500.000           | 2.4  |
| Decachlorobiphenyl   | 8.708 | 8.607                | 8.807 | 45.230             | 50.000            | -9.5 |
| Tetrachloro-m-xylene | 3.694 | 3.594                | 3.794 | 52.140             | 50.000            | 4.3  |